

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070621\
 Data File : PP037051.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Jul 2021 1:35
 Operator : DD\AJ
 Sample : PP070621ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP070621

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 07 08:36:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.958	3.949	2338757	1416249	51.458	51.756
2)	SA Decachlor...	10.989	9.244	2304974	1394218	50.070	50.089
Target Compounds							
3)	L1 AR-1016-1	6.279	5.201	877106	529514	504.748	512.399
4)	L1 AR-1016-2	6.303	5.221	1253913	728258	501.547	508.649
5)	L1 AR-1016-3	6.368	5.411	795833	403651	499.678	516.639
6)	L1 AR-1016-4	6.475	5.464	666827	310485	505.713	514.398
7)	L1 AR-1016-5	6.790	5.692	655972	402081	518.551	520.700
8)	L2 AR-1221-1	5.202	4.204	84721	54176	159.739	177.384
9)	L2 AR-1221-2	5.307	4.303	126404	75831	312.006	322.120
10)	L2 AR-1221-3	5.394	4.392	477254	285990	391.693	384.507
11)	L3 AR-1232-1	5.394	4.392	477254	285990	429.528	421.212
12)	L3 AR-1232-2	5.983	5.221	657484	728258	1102.655	1163.516
13)	L3 AR-1232-3	6.303	5.411	1253913	403651	1145.904	1215.768
14)	L3 AR-1232-4	6.475	5.509	666827	373991	1148.564	1356.462
15)	L3 AR-1232-5	6.573	5.692	528345	402081	1373.799	1352.577
16)	L4 AR-1242-1	6.279	5.201	877106	529514	688.095	702.554
17)	L4 AR-1242-2	6.303	5.221	1253913	728258	691.578	707.754
18)	L4 AR-1242-3	6.368	5.411	795833	403651	691.929	708.884
19)	L4 AR-1242-4	6.475	5.509	666827	373991	696.568	725.664
20)	L4 AR-1242-5	7.258	6.072	104014	286292	100.017	420.599 #
21)	L5 AR-1248-1	6.279	5.201	877106	529514	835.678	839.077
22)	L5 AR-1248-2	6.573	5.464	528345	310485	355.895	378.516
23)	L5 AR-1248-3	6.790	5.509	655972	373991	397.125	437.020
24)	L5 AR-1248-4	7.219	5.692	125695	402081	66.501	395.099 #
25)	L5 AR-1248-5	7.258	6.114	104014	49306	55.575	46.482
26)	L6 AR-1254-1	7.188	6.072	443908	286292	247.097	192.845
27)	L6 AR-1254-2	7.417	6.232	492749	260118	177.141	200.244
28)	L6 AR-1254-3	7.799	6.652	309717	155634	100.459	72.930 #
29)	L6 AR-1254-4	8.094	6.891	211170	80251	92.967	59.150 #
30)	L6 AR-1254-5	8.523	7.321	1579109	965070	631.039	514.803
31)	L7 AR-1260-1	7.966	6.789	1093677	696076	496.802	503.741
32)	L7 AR-1260-2	8.231	6.988	1332969	840614	479.180	500.385
33)	L7 AR-1260-3	8.597	7.142	1022120	800571	505.705	467.240
34)	L7 AR-1260-4	8.828	7.625	1213105	681696	493.170	501.421
35)	L7 AR-1260-5	9.156	7.874	2312051	1622045	501.051	502.893
36)	L8 AR-1262-1	8.597	7.321	1022120	965070	350.942	929.390 #
37)	L8 AR-1262-2	9.156	7.874	2312051	1622045	440.245	466.286
38)	L8 AR-1262-3	9.470	8.161	524342	381168	136.846	270.725 #
39)	L8 AR-1262-4	9.544f	8.223	1047761	1235745	356.573	464.044 #
40)	L8 AR-1262-5	10.234	8.723	786081	472689	369.692	373.962
41)	L9 AR-1268-1	9.470	8.161	524342	381168	86.343	95.648

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070621\
 Data File : PP037051.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Jul 2021 1:35
 Operator : DD\AJ
 Sample : PP070621ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP070621

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 07 08:36:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.544f	8.223	1047761	1235745	182.296	333.881 #
43) L9	AR-1268-3	9.791	8.434	39302	21221	8.015	6.756
44) L9	AR-1268-4	10.234	8.723	786081	472689	343.611	344.906
45) L9	AR-1268-5	10.650	9.002	241387	140678	15.246	13.668



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070621\
Data File : PP037051.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Jul 2021 1:35
Operator : DD\AJ
Sample : PP070621ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP070621

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 07 08:36:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.958 min
Delta R.T.: -0.002 min
Response: 2338757
Conc: 51.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#1 Tetrachloro-m-xylene

R.T.: 3.949 min
Delta R.T.: 0.000 min
Response: 1416249
Conc: 51.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.989 min
Delta R.T.: -0.003 min
Response: 2304974
Conc: 50.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.244 min
Delta R.T.: -0.003 min
Response: 1394218
Conc: 50.09 ng/ml



#3 AR-1016-1


R.T.: 6.279 min
Delta R.T.: -0.002 min
Response: 877106
Conc: 504.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#3 AR-1016-1

R.T.: 5.201 min
Delta R.T.: -0.001 min
Response: 529514
Conc: 512.40 ng/ml




#4 AR-1016-2

R.T.: 6.303 min
Delta R.T.: -0.001 min
Response: 1253913
Conc: 501.55 ng/ml



#4 AR-1016-2


R.T.: 5.221 min
Delta R.T.: 0.000 min
Response: 728258
Conc: 508.65 ng/ml



#5 AR-1016-3


R.T.: 6.368 min
Delta R.T.: -0.002 min
Response: 795833
Conc: 499.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#5 AR-1016-3

R.T.: 5.411 min
Delta R.T.: -0.001 min
Response: 403651
Conc: 516.64 ng/ml




#6 AR-1016-4

R.T.: 6.475 min
Delta R.T.: -0.002 min
Response: 666827
Conc: 505.71 ng/ml



#6 AR-1016-4


R.T.: 5.464 min
Delta R.T.: -0.001 min
Response: 310485
Conc: 514.40 ng/ml



#7 AR-1016-5


R.T.: 6.790 min
Delta R.T.: -0.002 min
Response: 655972
Conc: 518.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#7 AR-1016-5

R.T.: 5.692 min
Delta R.T.: -0.001 min
Response: 402081
Conc: 520.70 ng/ml




#8 AR-1221-1

R.T.: 5.202 min
Delta R.T.: -0.008 min
Response: 84721
Conc: 159.74 ng/ml



#8 AR-1221-1


R.T.: 4.204 min
Delta R.T.: -0.003 min
Response: 54176
Conc: 177.38 ng/ml



#9 AR-1221-2


R.T.: 5.307 min
Delta R.T.: -0.002 min
Response: 126404
Conc: 312.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#9 AR-1221-2

R.T.: 4.303 min
Delta R.T.: -0.002 min
Response: 75831
Conc: 322.12 ng/ml




#10 AR-1221-3

R.T.: 5.394 min
Delta R.T.: -0.002 min
Response: 477254
Conc: 391.69 ng/ml



#10 AR-1221-3


R.T.: 4.392 min
Delta R.T.: -0.002 min
Response: 285990
Conc: 384.51 ng/ml



#11 AR-1232-1


R.T.: 5.394 min
Delta R.T.: -0.002 min
Response: 477254
Conc: 429.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#11 AR-1232-1

R.T.: 4.392 min
Delta R.T.: -0.001 min
Response: 285990
Conc: 421.21 ng/ml




#12 AR-1232-2

R.T.: 5.983 min
Delta R.T.: -0.002 min
Response: 657484
Conc: 1102.66 ng/ml



#12 AR-1232-2


R.T.: 5.221 min
Delta R.T.: -0.001 min
Response: 728258
Conc: 1163.52 ng/ml



#13 AR-1232-3


R.T.: 6.303 min
Delta R.T.: -0.002 min
Response: 1253913
Conc: 1145.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#13 AR-1232-3

R.T.: 5.411 min
Delta R.T.: -0.001 min
Response: 403651
Conc: 1215.77 ng/ml




#14 AR-1232-4

R.T.: 6.475 min
Delta R.T.: -0.003 min
Response: 666827
Conc: 1148.56 ng/ml



#14 AR-1232-4


R.T.: 5.509 min
Delta R.T.: -0.002 min
Response: 373991
Conc: 1356.46 ng/ml



#15 AR-1232-5


R.T.: 6.573 min
Delta R.T.: -0.002 min
Response: 528345
Conc: 1373.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#15 AR-1232-5

R.T.: 5.692 min
Delta R.T.: -0.002 min
Response: 402081
Conc: 1352.58 ng/ml




#16 AR-1242-1

R.T.: 6.279 min
Delta R.T.: -0.002 min
Response: 877106
Conc: 688.09 ng/ml



#16 AR-1242-1


R.T.: 5.201 min
Delta R.T.: -0.002 min
Response: 529514
Conc: 702.55 ng/ml



#17 AR-1242-2


R.T.: 6.303 min
Delta R.T.: -0.002 min
Response: 1253913
Conc: 691.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#17 AR-1242-2

R.T.: 5.221 min
Delta R.T.: -0.002 min
Response: 728258
Conc: 707.75 ng/ml




#18 AR-1242-3

R.T.: 6.368 min
Delta R.T.: -0.002 min
Response: 795833
Conc: 691.93 ng/ml



#18 AR-1242-3

R.T.: 5.411 min
Delta R.T.: -0.001 min
Response: 403651
Conc: 708.88 ng/ml



#19 AR-1242-4

R.T.: 6.475 min
Delta R.T.: -0.003 min
Response: 666827
Conc: 696.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621

#19 AR-1242-4


R.T.: 5.509 min
Delta R.T.: -0.001 min
Response: 373991
Conc: 725.66 ng/ml

#20 AR-1242-5

R.T.: 7.258 min
Delta R.T.: -0.002 min
Response: 104014
Conc: 100.02 ng/ml

#20 AR-1242-5


R.T.: 6.072 min
Delta R.T.: -0.002 min
Response: 286292
Conc: 420.60 ng/ml



#21 AR-1248-1


R.T.: 6.279 min
Delta R.T.: -0.001 min
Response: 877106
Conc: 835.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#21 AR-1248-1

R.T.: 5.201 min
Delta R.T.: 0.000 min
Response: 529514
Conc: 839.08 ng/ml




#22 AR-1248-2

R.T.: 6.573 min
Delta R.T.: -0.001 min
Response: 528345
Conc: 355.90 ng/ml



#22 AR-1248-2


R.T.: 5.464 min
Delta R.T.: 0.000 min
Response: 310485
Conc: 378.52 ng/ml



#23 AR-1248-3


R.T.: 6.790 min
Delta R.T.: -0.001 min
Response: 655972
Conc: 397.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#23 AR-1248-3

R.T.: 5.509 min
Delta R.T.: 0.000 min
Response: 373991
Conc: 437.02 ng/ml




#24 AR-1248-4

R.T.: 7.219 min
Delta R.T.: -0.001 min
Response: 125695
Conc: 66.50 ng/ml



#24 AR-1248-4


R.T.: 5.692 min
Delta R.T.: 0.000 min
Response: 402081
Conc: 395.10 ng/ml



#25 AR-1248-5


R.T.: 7.258 min
Delta R.T.: 0.000 min
Response: 104014
Conc: 55.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#25 AR-1248-5

R.T.: 6.114 min
Delta R.T.: 0.000 min
Response: 49306
Conc: 46.48 ng/ml




#26 AR-1254-1

R.T.: 7.188 min
Delta R.T.: -0.002 min
Response: 443908
Conc: 247.10 ng/ml



#26 AR-1254-1


R.T.: 6.072 min
Delta R.T.: -0.001 min
Response: 286292
Conc: 192.84 ng/ml



#27 AR-1254-2


R.T.: 7.417 min
Delta R.T.: -0.002 min
Response: 492749
Conc: 177.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#27 AR-1254-2

R.T.: 6.232 min
Delta R.T.: -0.002 min
Response: 260118
Conc: 200.24 ng/ml




#28 AR-1254-3

R.T.: 7.799 min
Delta R.T.: 0.000 min
Response: 309717
Conc: 100.46 ng/ml



#28 AR-1254-3


R.T.: 6.652 min
Delta R.T.: 0.000 min
Response: 155634
Conc: 72.93 ng/ml



#29 AR-1254-4


R.T.: 8.094 min
Delta R.T.: -0.002 min
Response: 211170
Conc: 92.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#29 AR-1254-4

R.T.: 6.891 min
Delta R.T.: -0.002 min
Response: 80251
Conc: 59.15 ng/ml




#30 AR-1254-5

R.T.: 8.523 min
Delta R.T.: -0.002 min
Response: 1579109
Conc: 631.04 ng/ml



#30 AR-1254-5


R.T.: 7.321 min
Delta R.T.: -0.002 min
Response: 965070
Conc: 514.80 ng/ml



#31 AR-1260-1


R.T.: 7.966 min
Delta R.T.: -0.002 min
Response: 1093677
Conc: 496.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#31 AR-1260-1

R.T.: 6.789 min
Delta R.T.: -0.002 min
Response: 696076
Conc: 503.74 ng/ml




#32 AR-1260-2

R.T.: 8.231 min
Delta R.T.: -0.003 min
Response: 1332969
Conc: 479.18 ng/ml



#32 AR-1260-2


R.T.: 6.988 min
Delta R.T.: -0.002 min
Response: 840614
Conc: 500.39 ng/ml



#33 AR-1260-3


R.T.: 8.597 min
Delta R.T.: -0.002 min
Response: 1022120
Conc: 505.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#33 AR-1260-3

R.T.: 7.142 min
Delta R.T.: -0.001 min
Response: 800571
Conc: 467.24 ng/ml




#34 AR-1260-4

R.T.: 8.828 min
Delta R.T.: -0.002 min
Response: 1213105
Conc: 493.17 ng/ml



#34 AR-1260-4


R.T.: 7.625 min
Delta R.T.: -0.002 min
Response: 681696
Conc: 501.42 ng/ml



#35 AR-1260-5


R.T.: 9.156 min
Delta R.T.: -0.003 min
Response: 2312051
Conc: 501.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#35 AR-1260-5

R.T.: 7.874 min
Delta R.T.: -0.002 min
Response: 1622045
Conc: 502.89 ng/ml




#36 AR-1262-1

R.T.: 8.597 min
Delta R.T.: -0.001 min
Response: 1022120
Conc: 350.94 ng/ml



#36 AR-1262-1


R.T.: 7.321 min
Delta R.T.: 0.000 min
Response: 965070
Conc: 929.39 ng/ml



#37 AR-1262-2


R.T.: 9.156 min
Delta R.T.: 0.000 min
Response: 2312051
Conc: 440.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#37 AR-1262-2

R.T.: 7.874 min
Delta R.T.: -0.001 min
Response: 1622045
Conc: 466.29 ng/ml




#38 AR-1262-3

R.T.: 9.470 min
Delta R.T.: 0.000 min
Response: 524342
Conc: 136.85 ng/ml



#38 AR-1262-3


R.T.: 8.161 min
Delta R.T.: 0.000 min
Response: 381168
Conc: 270.72 ng/ml



#39 AR-1262-4


R.T.: 9.544 min
Delta R.T.: -0.018 min
Response: 1047761
Conc: 356.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#39 AR-1262-4

R.T.: 8.223 min
Delta R.T.: -0.002 min
Response: 1235745
Conc: 464.04 ng/ml




#40 AR-1262-5

R.T.: 10.234 min
Delta R.T.: 0.000 min
Response: 786081
Conc: 369.69 ng/ml



#40 AR-1262-5


R.T.: 8.723 min
Delta R.T.: 0.000 min
Response: 472689
Conc: 373.96 ng/ml



#41 AR-1268-1


R.T.: 9.470 min
Delta R.T.: 0.000 min
Response: 524342
Conc: 86.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#41 AR-1268-1

R.T.: 8.161 min
Delta R.T.: -0.002 min
Response: 381168
Conc: 95.65 ng/ml




#42 AR-1268-2

R.T.: 9.544 min
Delta R.T.: -0.023 min
Response: 1047761
Conc: 182.30 ng/ml



#42 AR-1268-2


R.T.: 8.223 min
Delta R.T.: -0.004 min
Response: 1235745
Conc: 333.88 ng/ml



#43 AR-1268-3


R.T.: 9.791 min
Delta R.T.: -0.002 min
Response: 39302
Conc: 8.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621




#43 AR-1268-3

R.T.: 8.434 min
Delta R.T.: -0.002 min
Response: 21221
Conc: 6.76 ng/ml




#44 AR-1268-4

R.T.: 10.234 min
Delta R.T.: -0.003 min
Response: 786081
Conc: 343.61 ng/ml



#44 AR-1268-4


R.T.: 8.723 min
Delta R.T.: -0.001 min
Response: 472689
Conc: 344.91 ng/ml



#45 AR-1268-5

R.T.: 10.650 min
Delta R.T.: -0.002 min
Response: 241387
Conc: 15.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621



#45 AR-1268-5

R.T.: 9.002 min
Delta R.T.: -0.002 min
Response: 140678
Conc: 13.67 ng/ml