

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070621\
 Data File : PP037031.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Jul 2021 19:56
 Operator : DD\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 07 07:46:57 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 07:40:05 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.959	3.950	4428179	2785523	94.676	95.670
2) SA Decachlor...	10.990	9.247	4159360	2627188	80.838	92.459
Target Compounds						
21) L5 AR-1248-1	6.280	5.202	917543	563037	881.538	914.867
22) L5 AR-1248-2	6.574	5.465	1288658	734030	891.523	899.832
23) L5 AR-1248-3	6.791	5.509	1455393	769836	891.341	904.177
24) L5 AR-1248-4	7.220	5.692	1650191	931766	895.209	914.417
25) L5 AR-1248-5	7.259	6.115	1635827	967738	897.156	917.775



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070621\
Data File : PP037031.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Jul 2021 19:56
Operator : DD\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 07 07:46:57 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 07:40:05 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.959 min
Delta R.T.: -0.002 min
Response: 4428179
Conc: 94.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.950 min
Delta R.T.: -0.001 min
Response: 2785523
Conc: 95.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.990 min
Delta R.T.: -0.010 min
Response: 4159360
Conc: 80.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.247 min
Delta R.T.: 0.000 min
Response: 2627188
Conc: 92.46 ng/ml



#21 AR-1248-1


R.T.: 6.280 min
Delta R.T.: -0.001 min
Response: 917543
Conc: 881.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 5.202 min
Delta R.T.: -0.002 min
Response: 563037
Conc: 914.87 ng/ml




#22 AR-1248-2

R.T.: 6.574 min
Delta R.T.: 0.000 min
Response: 1288658
Conc: 891.52 ng/ml



#22 AR-1248-2


R.T.: 5.465 min
Delta R.T.: -0.002 min
Response: 734030
Conc: 899.83 ng/ml



#23 AR-1248-3


R.T.: 6.791 min
Delta R.T.: -0.001 min
Response: 1455393
Conc: 891.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 5.509 min
Delta R.T.: -0.002 min
Response: 769836
Conc: 904.18 ng/ml




#24 AR-1248-4

R.T.: 7.220 min
Delta R.T.: 0.000 min
Response: 1650191
Conc: 895.21 ng/ml



#24 AR-1248-4


R.T.: 5.692 min
Delta R.T.: -0.002 min
Response: 931766
Conc: 914.42 ng/ml



#25 AR-1248-5

R.T.: 7.259 min
Delta R.T.: -0.002 min
Response: 1635827
Conc: 897.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 6.115 min
Delta R.T.: -0.002 min
Response: 967738
Conc: 917.78 ng/ml