

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070621\
 Data File : PP037054.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Jul 2021 2:26
 Operator : DD\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP070621AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 07 08:37:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.958	3.949	2356381	1439768	51.846	52.615
2) SA Decachlor...	10.987	9.244	2358947	1441201	51.242	51.777
Target Compounds						
26) L6 AR-1254-1	7.187	6.071	890605	736333	495.746	495.990
27) L6 AR-1254-2	7.417	6.232	1365997	644242	491.070	495.951
28) L6 AR-1254-3	7.799	6.650	1497590	1066055	485.755	499.551
29) L6 AR-1254-4	8.095	6.891	1124431	685817	495.030	505.493
30) L6 AR-1254-5	8.523	7.321	1234013	948333	493.133	505.875



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070621\
Data File : PP037054.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Jul 2021 2:26
Operator : DD\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP070621AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 07 08:37:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.958 min
Delta R.T.: -0.002 min
Response: 2356381
Conc: 51.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.949 min
Delta R.T.: 0.000 min
Response: 1439768
Conc: 52.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.987 min
Delta R.T.: -0.005 min
Response: 2358947
Conc: 51.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.244 min
Delta R.T.: -0.003 min
Response: 1441201
Conc: 51.78 ng/ml



#26 AR-1254-1


R.T.: 7.187 min
Delta R.T.: -0.002 min
Response: 890605
Conc: 495.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621AR1254




#26 AR-1254-1

R.T.: 6.071 min
Delta R.T.: -0.002 min
Response: 736333
Conc: 495.99 ng/ml




#27 AR-1254-2

R.T.: 7.417 min
Delta R.T.: -0.002 min
Response: 1365997
Conc: 491.07 ng/ml



#27 AR-1254-2


R.T.: 6.232 min
Delta R.T.: -0.002 min
Response: 644242
Conc: 495.95 ng/ml



#28 AR-1254-3


R.T.: 7.799 min
Delta R.T.: -0.001 min
Response: 1497590
Conc: 485.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621AR1254




#28 AR-1254-3

R.T.: 6.650 min
Delta R.T.: -0.002 min
Response: 1066055
Conc: 499.55 ng/ml




#29 AR-1254-4

R.T.: 8.095 min
Delta R.T.: -0.002 min
Response: 1124431
Conc: 495.03 ng/ml



#29 AR-1254-4


R.T.: 6.891 min
Delta R.T.: -0.002 min
Response: 685817
Conc: 505.49 ng/ml



#30 AR-1254-5

R.T.: 8.523 min
Delta R.T.: -0.002 min
Response: 1234013
Conc: 493.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070621AR1254



#30 AR-1254-5

R.T.: 7.321 min
Delta R.T.: -0.002 min
Response: 948333
Conc: 505.88 ng/ml