

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070622\
 Data File : PP049229.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jul 2022 20:12
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 07 03:15:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 09:37:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.345	3.595	75009562	128.5E6	52.765	50.165
2) SA Decachlor...	10.094	8.567	66900978	51782250	54.542	45.938
Target Compounds						
16) L4 AR-1242-1	5.523	4.670	21805673	32921299	538.100	457.751
17) L4 AR-1242-2	5.545	4.689	30835735	46022311	535.416	454.401
18) L4 AR-1242-3	5.607	4.864	19396708	24543802	535.749	440.288
19) L4 AR-1242-4	5.707	4.946	16178807	24265523	549.942	438.329
20) L4 AR-1242-5	6.447	5.466	17039154	29954641	510.608	430.254



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070622\
Data File : PP049229.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jul 2022 20:12
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 07 03:15:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 29 09:37:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.345 min
Delta R.T.: -0.001 min
Response: 75009562
Conc: 52.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.595 min
Delta R.T.: 0.000 min
Response: 128455598
Conc: 50.17 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.094 min
Delta R.T.: 0.000 min
Response: 66900978
Conc: 54.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.567 min
Delta R.T.: 0.000 min
Response: 51782250
Conc: 45.94 ng/ml



#16 AR-1242-1


R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 21805673
Conc: 538.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.670 min
Delta R.T.: 0.000 min
Response: 32921299
Conc: 457.75 ng/ml




#17 AR-1242-2

R.T.: 5.545 min
Delta R.T.: 0.000 min
Response: 30835735
Conc: 535.42 ng/ml



#17 AR-1242-2


R.T.: 4.689 min
Delta R.T.: 0.000 min
Response: 46022311
Conc: 454.40 ng/ml



#18 AR-1242-3


R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 19396708
Conc: 535.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 24543802
Conc: 440.29 ng/ml




#19 AR-1242-4

R.T.: 5.707 min
Delta R.T.: 0.000 min
Response: 16178807
Conc: 549.94 ng/ml



#19 AR-1242-4


R.T.: 4.946 min
Delta R.T.: 0.000 min
Response: 24265523
Conc: 438.33 ng/ml



#20 AR-1242-5

R.T.: 6.447 min
Delta R.T.: 0.000 min
Response: 17039154
Conc: 510.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.466 min
Delta R.T.: 0.000 min
Response: 29954641
Conc: 430.25 ng/ml