

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070722\
 Data File : PP049288.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jul 2022 16:29
 Operator : YP\AJ
 Sample : N3615-09
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 S3-02-S

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/08/2022
 Supervised By :Ankita Jodhani 07/08/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 00:01:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 09:37:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.597	31282790	56001293	22.006	21.870
2) SA Decachlor...	10.093	8.567	24203379	19534522	19.732	17.330

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070722\
Data File : PP049288.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2022 16:29
Operator : YP\AJ
Sample : N3615-09
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
S3-02-S


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2022
Supervised By :Ankita Jodhani 07/08/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 00:01:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 29 09:37:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.347 min
Delta R.T.: 0.001 min
Response: 31282790
Conc: 22.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
S3-02-S


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2022
Supervised By :Ankita Jodhani 07/08/2022




#1 Tetrachloro-m-xylene

R.T.: 3.597 min
Delta R.T.: 0.003 min
Response: 56001293
Conc: 21.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.093 min
Delta R.T.: 0.000 min
Response: 24203379
Conc: 19.73 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.567 min
Delta R.T.: 0.000 min
Response: 19534522
Conc: 17.33 ng/ml