

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071025\
 Data File : PP073673.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2025 03:06
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 03:33:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 08 08:35:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.487	3.777	71903004	95926040	52.502	52.069
2) SA Decachlor...	10.174	8.777	59263025	72492279	54.317	54.788
Target Compounds						
26) L6 AR-1254-1	6.485	5.655	27324784	56978995	509.890	491.071
27) L6 AR-1254-2	6.702	5.803	41455578	49194291	497.732	488.894
28) L6 AR-1254-3	7.065	6.205	45680079	75655508	517.143	489.849
29) L6 AR-1254-4	7.347	6.434	38974657	47849101	495.438	506.023
30) L6 AR-1254-5	7.762	6.849	37206718	64887739	504.979	486.251



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071025\
Data File : PP073673.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 03:06
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 03:33:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 08 08:35:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.487 min
Delta R.T.: -0.001 min
Response: 71903004
Conc: 52.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.777 min
Delta R.T.: -0.001 min
Response: 95926040
Conc: 52.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.174 min
Delta R.T.: -0.002 min
Response: 59263025
Conc: 54.32 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.777 min
Delta R.T.: -0.003 min
Response: 72492279
Conc: 54.79 ng/ml



#26 AR-1254-1


R.T.: 6.485 min
Delta R.T.: 0.000 min
Response: 27324784
Conc: 509.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.655 min
Delta R.T.: -0.003 min
Response: 56978995
Conc: 491.07 ng/ml




#27 AR-1254-2

R.T.: 6.702 min
Delta R.T.: 0.000 min
Response: 41455578
Conc: 497.73 ng/ml



#27 AR-1254-2


R.T.: 5.803 min
Delta R.T.: -0.003 min
Response: 49194291
Conc: 488.89 ng/ml



#28 AR-1254-3


R.T.: 7.065 min
Delta R.T.: 0.000 min
Response: 45680079
Conc: 517.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.205 min
Delta R.T.: -0.004 min
Response: 75655508
Conc: 489.85 ng/ml




#29 AR-1254-4

R.T.: 7.347 min
Delta R.T.: 0.000 min
Response: 38974657
Conc: 495.44 ng/ml



#29 AR-1254-4


R.T.: 6.434 min
Delta R.T.: -0.002 min
Response: 47849101
Conc: 506.02 ng/ml



#30 AR-1254-5

R.T.: 7.762 min
Delta R.T.: 0.000 min
Response: 37206718
Conc: 504.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.849 min
Delta R.T.: -0.003 min
Response: 64887739
Conc: 486.25 ng/ml