

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071122\
 Data File : PP049334.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jul 2022 11:15
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 11 23:55:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 09:37:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.344	3.595	79002085	131.3E6	55.574	51.278
2) SA Decachlor...	10.094	8.567	71193746	56252281	58.042	49.904
Target Compounds						
3) L1 AR-1016-1	5.521	4.669	30067201	43257195	571.919	473.951
4) L1 AR-1016-2	5.545	4.688	43150614	61704222	572.593	479.475
5) L1 AR-1016-3	5.606	4.863	26287018	32508824	539.977	439.323
6) L1 AR-1016-4	5.707	4.904	22004477	26117945	566.889	510.531
7) L1 AR-1016-5	6.001	5.117	21724369	33682285	569.399	451.514
31) L7 AR-1260-1	7.129	6.144	40349943	49470031	554.654	429.414
32) L7 AR-1260-2	7.388	6.332	47901328	57734702	569.142	430.436
33) L7 AR-1260-3	7.746	6.485	32336339	52903211	577.993	434.584
34) L7 AR-1260-4	7.974	6.955	39383858	36877495	563.162	437.083
35) L7 AR-1260-5	8.295	7.196	77413695	82419710	564.972	437.187



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071122\
Data File : PP049334.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jul 2022 11:15
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 11 23:55:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 29 09:37:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.344 min
Delta R.T.: -0.002 min
Response: 79002085
Conc: 55.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.595 min
Delta R.T.: 0.000 min
Response: 131304467
Conc: 51.28 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.094 min
Delta R.T.: 0.000 min
Response: 71193746
Conc: 58.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.567 min
Delta R.T.: 0.000 min
Response: 56252281
Conc: 49.90 ng/ml



#3 AR-1016-1


R.T.: 5.521 min
Delta R.T.: -0.001 min
Response: 30067201
Conc: 571.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.669 min
Delta R.T.: 0.000 min
Response: 43257195
Conc: 473.95 ng/ml




#4 AR-1016-2

R.T.: 5.545 min
Delta R.T.: 0.000 min
Response: 43150614
Conc: 572.59 ng/ml



#4 AR-1016-2


R.T.: 4.688 min
Delta R.T.: 0.000 min
Response: 61704222
Conc: 479.47 ng/ml



#5 AR-1016-3


R.T.: 5.606 min
Delta R.T.: -0.001 min
Response: 26287018
Conc: 539.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.863 min
Delta R.T.: -0.001 min
Response: 32508824
Conc: 439.32 ng/ml




#6 AR-1016-4

R.T.: 5.707 min
Delta R.T.: -0.001 min
Response: 22004477
Conc: 566.89 ng/ml



#6 AR-1016-4


R.T.: 4.904 min
Delta R.T.: -0.001 min
Response: 26117945
Conc: 510.53 ng/ml



#7 AR-1016-5


R.T.: 6.001 min
Delta R.T.: 0.000 min
Response: 21724369
Conc: 569.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 33682285
Conc: 451.51 ng/ml




#31 AR-1260-1

R.T.: 7.129 min
Delta R.T.: -0.001 min
Response: 40349943
Conc: 554.65 ng/ml



#31 AR-1260-1


R.T.: 6.144 min
Delta R.T.: 0.000 min
Response: 49470031
Conc: 429.41 ng/ml



#32 AR-1260-2


R.T.: 7.388 min
Delta R.T.: 0.000 min
Response: 47901328
Conc: 569.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.332 min
Delta R.T.: 0.000 min
Response: 57734702
Conc: 430.44 ng/ml




#33 AR-1260-3

R.T.: 7.746 min
Delta R.T.: -0.001 min
Response: 32336339
Conc: 577.99 ng/ml



#33 AR-1260-3


R.T.: 6.485 min
Delta R.T.: -0.001 min
Response: 52903211
Conc: 434.58 ng/ml



#34 AR-1260-4


R.T.: 7.974 min
Delta R.T.: -0.001 min
Response: 39383858
Conc: 563.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.955 min
Delta R.T.: -0.002 min
Response: 36877495
Conc: 437.08 ng/ml



#35 AR-1260-5

R.T.: 8.295 min
Delta R.T.: 0.000 min
Response: 77413695
Conc: 564.97 ng/ml



#35 AR-1260-5

R.T.: 7.196 min
Delta R.T.: -0.001 min
Response: 82419710
Conc: 437.19 ng/ml