

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071222\
 Data File : PP049391.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jul 2022 14:36
 Operator : YP\AJ
 Sample : PB146168BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB146168BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 13 04:57:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 09:37:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.345	3.596	32728780	54593107	23.023	21.320
2) SA Decachlor...	10.098	8.568	31119899	26327069	25.371	23.356
Target Compounds						
3) L1 AR-1016-1	5.523	4.671	28087081	42346286	534.255	463.970
4) L1 AR-1016-2	5.546	4.689	40373635	60488433	535.743	470.028
5) L1 AR-1016-3	5.608	4.864	25029673	31854181	514.150	430.476
6) L1 AR-1016-4	5.708	4.905	20781101	25056827	535.371	489.789
7) L1 AR-1016-5	6.002	5.117	20993089	31373500	550.232	420.565
31) L7 AR-1260-1	7.132	6.146	40070001	50240028	550.806	436.098
32) L7 AR-1260-2	7.390	6.333	47951090	58313618	569.733	434.752
33) L7 AR-1260-3	7.748	6.486	30303184	53858386	541.652	442.431
34) L7 AR-1260-4	7.977	6.957	37330471	36217075	533.800	429.256
35) L7 AR-1260-5	8.297	7.198	72860000	81207429	531.739	430.756



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071222\
Data File : PP049391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jul 2022 14:36
Operator : YP\AJ
Sample : PB146168BS
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB146168BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 13 04:57:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 29 09:37:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.345 min
Delta R.T.: 0.000 min
Response: 32728780
Conc: 23.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146168BS




#1 Tetrachloro-m-xylene

R.T.: 3.596 min
Delta R.T.: 0.002 min
Response: 54593107
Conc: 21.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.098 min
Delta R.T.: 0.003 min
Response: 31119899
Conc: 25.37 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.568 min
Delta R.T.: 0.001 min
Response: 26327069
Conc: 23.36 ng/ml



#3 AR-1016-1


R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 28087081
Conc: 534.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146168BS




#3 AR-1016-1

R.T.: 4.671 min
Delta R.T.: 0.000 min
Response: 42346286
Conc: 463.97 ng/ml




#4 AR-1016-2

R.T.: 5.546 min
Delta R.T.: 0.000 min
Response: 40373635
Conc: 535.74 ng/ml



#4 AR-1016-2


R.T.: 4.689 min
Delta R.T.: 0.000 min
Response: 60488433
Conc: 470.03 ng/ml



#5 AR-1016-3


R.T.: 5.608 min
Delta R.T.: 0.000 min
Response: 25029673
Conc: 514.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146168BS




#5 AR-1016-3

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 31854181
Conc: 430.48 ng/ml




#6 AR-1016-4

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 20781101
Conc: 535.37 ng/ml



#6 AR-1016-4


R.T.: 4.905 min
Delta R.T.: 0.000 min
Response: 25056827
Conc: 489.79 ng/ml



#7 AR-1016-5


R.T.: 6.002 min
Delta R.T.: 0.000 min
Response: 20993089
Conc: 550.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146168BS




#7 AR-1016-5

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 31373500
Conc: 420.56 ng/ml




#31 AR-1260-1

R.T.: 7.132 min
Delta R.T.: 0.001 min
Response: 40070001
Conc: 550.81 ng/ml



#31 AR-1260-1


R.T.: 6.146 min
Delta R.T.: 0.000 min
Response: 50240028
Conc: 436.10 ng/ml



#32 AR-1260-2


R.T.: 7.390 min
Delta R.T.: 0.001 min
Response: 47951090
Conc: 569.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146168BS




#32 AR-1260-2

R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 58313618
Conc: 434.75 ng/ml




#33 AR-1260-3

R.T.: 7.748 min
Delta R.T.: 0.000 min
Response: 30303184
Conc: 541.65 ng/ml



#33 AR-1260-3


R.T.: 6.486 min
Delta R.T.: 0.000 min
Response: 53858386
Conc: 442.43 ng/ml



#34 AR-1260-4


R.T.: 7.977 min
Delta R.T.: 0.001 min
Response: 37330471
Conc: 533.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146168BS




#34 AR-1260-4

R.T.: 6.957 min
Delta R.T.: 0.000 min
Response: 36217075
Conc: 429.26 ng/ml



#35 AR-1260-5

R.T.: 8.297 min
Delta R.T.: 0.002 min
Response: 72860000
Conc: 531.74 ng/ml



#35 AR-1260-5

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 81207429
Conc: 430.76 ng/ml