

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD P\Data\PP071318\
 Data File : PP026737.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Jul 2018 21:54
 Operator : EI\MA
 Sample : J3762-02
 Misc : L00
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 01:24:27 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071318.M
 Quant Title : 8080.M
 QLast Update : Fri Jul 13 23:16:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.045	12.157	116.4E6	167.3E6	521.721	418.734
Target Compounds							
1) T	Dalapon	3.307	2.981	143.6E6	133.1E6	278.901	238.122
2) T	3,5-DICHL...	10.459	10.249	128.8E6	146.6E6	286.745	274.809
3) T	4-Nitroph...	11.618	11.297	128.8E6	70966050	273.578	297.526
5) T	DICAMBA	12.368	12.468	340.2E6	471.0E6	275.144	295.561
6) T	MCPD	12.719	0.000	25448866	0	64.534	N.D. #
7) T	MCPA	12.950	0.000	14321798	0	24.014	N.D. #
8) T	DICHLORPROP	13.564	13.582	88519305	126.2E6	294.290	326.543
9) T	2,4-D	13.912	14.067	110.9E6	160.7E6	308.380	306.602
10) T	Pentachlo...	14.331	14.680	2061.0E6	2058.7E6	304.729	311.861
11) T	2,4,5-TP ...	15.193	15.250	743.3E6	780.5E6	294.430	301.961
12) T	2,4,5-T	15.580	15.887	644.8E6	640.7E6	282.503	291.215
13) T	2,4-DB	16.343	16.491	79744739	87970815	406.149	304.342 #
14) T	DINOSEB	17.834	16.885	469.5E6	429.7E6	255.509	240.838
15) T	Picloram	17.589	18.257	896.4E6	856.9E6	235.121	235.902
16) T	DCPA	18.197	18.144	724.3E6	702.7E6	198.043	218.877



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD P\Data\PP071318\
Data File : PP026737.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Jul 2018 21:54
Operator : EI\MA
Sample : J3762-02
Misc : L00
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 01:24:27 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071318.M
Quant Title : 8080.M
QLast Update : Fri Jul 13 23:16:25 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.307 min
Delta R.T.: -0.005 min
Response: 143625209
Conc: 278.90 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Dalapon

R.T.: 2.981 min
Delta R.T.: -0.002 min
Response: 133124397
Conc: 238.12 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.459 min
Delta R.T.: -0.002 min
Response: 128846970
Conc: 286.75 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 10.249 min
Delta R.T.: -0.001 min
Response: 146618921
Conc: 274.81 ng/ml



#3 4-Nitrophenol


R.T.: 11.618 min
Delta R.T.: -0.002 min
Response: 128788852
Conc: 273.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 4-Nitrophenol

R.T.: 11.297 min
Delta R.T.: -0.003 min
Response: 70966050
Conc: 297.53 ng/ml




#4 2,4-DCAA

R.T.: 12.045 min
Delta R.T.: -0.001 min
Response: 116423754
Conc: 521.72 ng/ml



#4 2,4-DCAA


R.T.: 12.157 min
Delta R.T.: -0.001 min
Response: 167295146
Conc: 418.73 ng/ml



#5 DICAMBA


R.T.: 12.368 min
Delta R.T.: 0.000 min
Response: 340214259
Conc: 275.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 DICAMBA

R.T.: 12.468 min
Delta R.T.: -0.002 min
Response: 471016643
Conc: 295.56 ng/ml




#6 MCP

R.T.: 12.719 min
Delta R.T.: 0.011 min
Response: 25448866
Conc: 64.53 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 12.657 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 12.950 min
Delta R.T.: 0.000 min
Response: 14321798
Conc: 24.01 ug/ml

Instrument :
ECD_P
ClientSampleId :




#7 MCPA

R.T.: 0.000 min
Exp R.T.: 13.044 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 13.564 min
Delta R.T.: -0.001 min
Response: 88519305
Conc: 294.29 ng/ml



#8 DICHLORPROP


R.T.: 13.582 min
Delta R.T.: -0.001 min
Response: 126176642
Conc: 326.54 ng/ml



#9 2,4-D


R.T.: 13.912 min
Delta R.T.: -0.002 min
Response: 110926601
Conc: 308.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 2,4-D

R.T.: 14.067 min
Delta R.T.: -0.002 min
Response: 160738788
Conc: 306.60 ng/ml




#10 Pentachlorophenol

R.T.: 14.331 min
Delta R.T.: 0.000 min
Response: 2061046537
Conc: 304.73 ng/ml



#10 Pentachlorophenol


R.T.: 14.680 min
Delta R.T.: 0.000 min
Response: 2058710692
Conc: 311.86 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 15.193 min
Delta R.T.: -0.001 min
Response: 743254605
Conc: 294.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 15.250 min
Delta R.T.: -0.001 min
Response: 780490030
Conc: 301.96 ng/ml




#12 2,4,5-T

R.T.: 15.580 min
Delta R.T.: -0.002 min
Response: 644814575
Conc: 282.50 ng/ml



#12 2,4,5-T


R.T.: 15.887 min
Delta R.T.: -0.002 min
Response: 640692743
Conc: 291.22 ng/ml



#13 2,4-DB


R.T.: 16.343 min
Delta R.T.: -0.002 min
Response: 79744739
Conc: 406.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 2,4-DB

R.T.: 16.491 min
Delta R.T.: -0.002 min
Response: 87970815
Conc: 304.34 ng/ml




#14 DINOSEB

R.T.: 17.834 min
Delta R.T.: -0.002 min
Response: 469530231
Conc: 255.51 ng/ml



#14 DINOSEB


R.T.: 16.885 min
Delta R.T.: -0.001 min
Response: 429702802
Conc: 240.84 ng/ml



#15 Picloram


R.T.: 17.589 min
Delta R.T.: -0.002 min
Response: 896354464
Conc: 235.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 Picloram

R.T.: 18.257 min
Delta R.T.: -0.001 min
Response: 856894187
Conc: 235.90 ng/ml



#16 DCPA

R.T.: 18.197 min
Delta R.T.: -0.003 min
Response: 724342423
Conc: 198.04 ng/ml



#16 DCPA

R.T.: 18.144 min
Delta R.T.: -0.001 min
Response: 702695532
Conc: 218.88 ng/ml