

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD P\methods\
 Method File : PP071318.M
 Title : 8080.M
 Last Update : Fri Jul 13 23:16:25 2018
 Response Via : Initial Calibration

Calibration Files

750 =PP026728.D 1000 =PP026729.D 1500 =PP026730.D
 500 =PP026727.D 200 =PP026726.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.122	5.102	4.971	5.189	5.365	5.150	E5	2.80
2) T	3,5-DICHLOROBENZOI...	4.485	4.479	4.235	4.536	4.731	4.493	E5	3.94
3) T	4-Nitrophenol	4.742	4.730	4.612	4.694	4.759	4.708	E5	1.25
4) S	2,4-DCAA	2.211	2.200	2.134	2.257	2.356	2.232	E5	3.68
5) T	DICAMBA	1.234	1.235	1.198	1.238	1.278	1.236	E6	2.30
6) T	MCPP	4.129	4.356	4.375	3.795	3.064	3.943	E5	13.81
7) T	MCPA	6.344	6.380	5.912	6.159	5.025	5.964	E5	9.34
8) T	DICHLORPROP	2.998	2.971	2.761	3.031	3.278	3.008	E5	6.13
9) T	2,4-D	3.546	3.531	3.406	3.638	3.865	3.597	E5	4.75
10) T	Pentachlorophenol	6.803	6.609	6.102	7.047	7.256	6.764	E6	6.55
11) T	2,4,5-TP (SILVEX)	2.528	2.499	2.379	2.573	2.642	2.524	E6	3.85
12) T	2,4,5-T	2.289	2.267	2.169	2.325	2.363	2.283	E6	3.20
13) T	2,4-DB	1.972	1.955	1.896	1.964	2.030	1.963	E5	2.43
14) T	DINOSEB	1.843	1.831	1.760	1.858	1.897	1.838	E6	2.71
15) T	Picloram	3.854	3.851	3.688	3.848	3.820	3.812	E6	1.85
16) T	DCPA	3.676	3.628	3.405	3.765	3.814	3.658	E6	4.35

Signal #2 Calibration Files

750 =PP026728.D 1000 =PP026729.D 1500 =PP026730.D
 500 =PP026727.D 200 =PP026726.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.479	5.477	5.396	5.620	5.980	5.591	E5	4.15
2) T	3,5-DICHLOROBENZOI...	5.296	5.287	5.056	5.386	5.651	5.335	E5	4.02
3) T	4-Nitrophenol	2.407	2.430	2.348	2.399	2.342	2.385	E5	1.61
4) S	2,4-DCAA	3.952	3.938	3.778	4.028	4.280	3.995	E5	4.59
5) T	DICAMBA	1.588	1.577	1.494	1.615	1.693	1.594	E6	4.49
6) T	MCPP	9.519	9.836	9.751	9.147	7.840	9.219	E5	8.85
7) T	MCPA	1.517	1.545	1.512	1.492	1.115	1.436	E6	12.58
8) T	DICHLORPROP	3.819	3.806	3.645	3.892	4.158	3.864	E5	4.85
9) T	2,4-D	5.206	5.181	4.944	5.309	5.572	5.243	E5	4.34
10) T	Pentachlorophenol	6.562	6.279	5.616	6.941	7.609	6.601	E6	11.25
11) T	2,4,5-TP (SILVEX)	2.579	2.525	2.351	2.659	2.810	2.585	E6	6.55
12) T	2,4,5-T	2.195	2.157	2.015	2.257	2.376	2.200	E6	6.02
13) T	2,4-DB	2.863	2.850	2.736	2.893	3.112	2.891	E5	4.75
14) T	DINOSEB	1.784	1.751	1.652	1.828	1.905	1.784	E6	5.27
15) T	Picloram	3.667	3.600	3.344	3.736	3.816	3.632	E6	4.96
16) T	DCPA	3.203	3.094	2.842	3.340	3.574	3.210	E6	8.51

(#) = Out of Range