

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071321\
 Data File : PP037259.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Jul 2021 12:29
 Operator : DD\AJ
 Sample : M3017-01 10X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 P001-B46C06-01

Manual Integrations
 APPROVED

mohammad
 7/14/2021 12:22:07 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 13 14:13:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.955	3.946	105502	62310	2.321	2.277
2)	SA Decachlor...	10.982	9.239	119805	65008	2.602	2.335
Target Compounds							
26)	L6 AR-1254-1	7.184	6.067	304895	215060	169.717m	144.863
27)	L6 AR-1254-2	7.414	6.228	570930	266187	205.247	204.917
28)	L6 AR-1254-3	7.796	6.647	633976	421110	205.635	197.331
29)	L6 AR-1254-4	8.092	6.887	588308	333049	259.002	245.479
30)	L6 AR-1254-5	8.520	7.317	699923	504654	279.701	269.201

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071321\
Data File : PP037259.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Jul 2021 12:29
Operator : DD\AJ
Sample : M3017-01 10X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
P001-B46C06-01

Manual Integrations
APPROVED

mohammad
7/14/2021 12:22:07 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 13 14:13:00 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

