

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071421\
 Data File : PP037276.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Jul 2021 15:21
 Operator : DD\AJ
 Sample : M3026-05
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 18:07:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.957	3.948	667157	396738	14.679	14.498
2)	SA Decachlor...	10.988	9.241	493434	240971	10.719	8.657
Target Compounds							
20)	L4 AR-1242-5	0.000	6.069	0	13606	N.D.	19.989 #
26)	L6 AR-1254-1	7.189	6.069	14080	13606	7.838	9.165
27)	L6 AR-1254-2	7.421	0.000	30239	0	10.871	N.D. #
28)	L6 AR-1254-3	7.799	6.648	41144	29949	13.346	14.034
29)	L6 AR-1254-4	8.092	0.000	12804	0	5.637	N.D. #
30)	L6 AR-1254-5	8.524	7.318	24113	105136	9.636	56.083 #
31)	L7 AR-1260-1	7.966	6.787	15264	36700	6.934	26.559 #
32)	L7 AR-1260-2	8.229	6.985	57428	15843	20.644	9.431 #
33)	L7 AR-1260-3	8.597	7.136	14816	23957	7.330	13.982 #
34)	L7 AR-1260-4	8.824	0.000	47892	0	19.470	N.D. #
35)	L7 AR-1260-5	9.156	7.872	32518	20454	7.047	6.341
36)	L8 AR-1262-1	8.597	7.318	14816	105136	5.087	101.249 #
37)	L8 AR-1262-2	9.156	7.872	32518	20454	6.192	5.880
38)	L8 AR-1262-3	9.488f	0.000	26251	0	6.851	N.D. #
39)	L8 AR-1262-4	0.000	8.219	0	17755	N.D.	6.667 #
40)	L8 AR-1262-5	10.234	0.000	14980	0	7.045	N.D. #
41)	L9 AR-1268-1	9.488f	0.000	26251	0	4.323	N.D. #
42)	L9 AR-1268-2	0.000	8.219	0	17755	N.D.	4.797 #
44)	L9 AR-1268-4	10.234	0.000	14980	0	6.548	N.D. #
45)	L9 AR-1268-5	10.651	8.999	28243	10681	1.784	1.038 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071421\
Data File : PP037276.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Jul 2021 15:21
Operator : DD\AJ
Sample : M3026-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 18:07:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.957 min
Delta R.T.: -0.003 min
Response: 667157
Conc: 14.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.948 min
Delta R.T.: -0.002 min
Response: 396738
Conc: 14.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.988 min
Delta R.T.: -0.004 min
Response: 493434
Conc: 10.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.241 min
Delta R.T.: -0.007 min
Response: 240971
Conc: 8.66 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 7.260 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#20 AR-1242-5

R.T.: 6.069 min
Delta R.T.: -0.004 min
Response: 13606
Conc: 19.99 ng/ml




#26 AR-1254-1

R.T.: 7.189 min
Delta R.T.: 0.000 min
Response: 14080
Conc: 7.84 ng/ml



#26 AR-1254-1


R.T.: 6.069 min
Delta R.T.: -0.003 min
Response: 13606
Conc: 9.16 ng/ml



#27 AR-1254-2


R.T.: 7.421 min
Delta R.T.: 0.003 min
Response: 30239
Conc: 10.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 6.234 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.799 min
Delta R.T.: 0.000 min
Response: 41144
Conc: 13.35 ng/ml



#28 AR-1254-3


R.T.: 6.648 min
Delta R.T.: -0.005 min
Response: 29949
Conc: 14.03 ng/ml



#29 AR-1254-4


R.T.: 8.092 min
Delta R.T.: -0.004 min
Response: 12804
Conc: 5.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.893 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 8.524 min
Delta R.T.: -0.001 min
Response: 24113
Conc: 9.64 ng/ml



#30 AR-1254-5


R.T.: 7.318 min
Delta R.T.: -0.004 min
Response: 105136
Conc: 56.08 ng/ml



#31 AR-1260-1


R.T.: 7.966 min
Delta R.T.: -0.001 min
Response: 15264
Conc: 6.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.787 min
Delta R.T.: -0.004 min
Response: 36700
Conc: 26.56 ng/ml




#32 AR-1260-2

R.T.: 8.229 min
Delta R.T.: -0.005 min
Response: 57428
Conc: 20.64 ng/ml



#32 AR-1260-2


R.T.: 6.985 min
Delta R.T.: -0.004 min
Response: 15843
Conc: 9.43 ng/ml



#33 AR-1260-3


R.T.: 8.597 min
Delta R.T.: -0.002 min
Response: 14816
Conc: 7.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.136 min
Delta R.T.: -0.007 min
Response: 23957
Conc: 13.98 ng/ml




#34 AR-1260-4

R.T.: 8.824 min
Delta R.T.: -0.005 min
Response: 47892
Conc: 19.47 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T.: 7.627 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 9.156 min
Delta R.T.: -0.002 min
Response: 32518
Conc: 7.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.872 min
Delta R.T.: -0.004 min
Response: 20454
Conc: 6.34 ng/ml




#36 AR-1262-1

R.T.: 8.597 min
Delta R.T.: 0.000 min
Response: 14816
Conc: 5.09 ng/ml



#36 AR-1262-1


R.T.: 7.318 min
Delta R.T.: -0.003 min
Response: 105136
Conc: 101.25 ng/ml



#37 AR-1262-2


R.T.: 9.156 min
Delta R.T.: 0.000 min
Response: 32518
Conc: 6.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.872 min
Delta R.T.: -0.004 min
Response: 20454
Conc: 5.88 ng/ml




#38 AR-1262-3

R.T.: 9.488 min
Delta R.T.: 0.019 min
Response: 26251
Conc: 6.85 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T.: 8.162 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 9.562 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.219 min
Delta R.T.: -0.006 min
Response: 17755
Conc: 6.67 ng/ml




#40 AR-1262-5

R.T.: 10.234 min
Delta R.T.: 0.000 min
Response: 14980
Conc: 7.04 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 8.724 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 9.488 min
Delta R.T.: 0.018 min
Response: 26251
Conc: 4.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 8.163 min
Response: 0
Conc: N.D.




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 9.567 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 8.219 min
Delta R.T.: -0.009 min
Response: 17755
Conc: 4.80 ng/ml



#44 AR-1268-4


R.T.: 10.234 min
Delta R.T.: -0.002 min
Response: 14980
Conc: 6.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 8.725 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 10.651 min
Delta R.T.: -0.001 min
Response: 28243
Conc: 1.78 ng/ml



#45 AR-1268-5

R.T.: 8.999 min
Delta R.T.: -0.004 min
Response: 10681
Conc: 1.04 ng/ml