

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071422\
 Data File : PP049439.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2022 21:51
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 05:22:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 05:21:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.352	3.601	112.9E6	198.5E6	73.126	72.822
2)	SA Decachlor...	10.100	8.569	111.5E6	89500198	72.308	71.862
Target Compounds							
21)	L5 AR-1248-1	5.529	4.675	25295713	40253119	714.484	710.485
22)	L5 AR-1248-2	5.801	4.909	35247884	52563187	706.445	705.772
23)	L5 AR-1248-3	6.008	4.950	39951060	54082985	708.650	703.232
24)	L5 AR-1248-4	6.414	5.121	45041663	64278321	709.929	706.324
25)	L5 AR-1248-5	6.452	5.511	43089621	59252859	713.148	708.985



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071422\
Data File : PP049439.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2022 21:51
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 05:22:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 05:21:07 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.352 min
Delta R.T.: 0.000 min
Response: 112922393
Conc: 73.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.601 min
Delta R.T.: 0.000 min
Response: 198494885
Conc: 72.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.100 min
Delta R.T.: -0.001 min
Response: 111472909
Conc: 72.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.569 min
Delta R.T.: 0.001 min
Response: 89500198
Conc: 71.86 ng/ml



#21 AR-1248-1


R.T.: 5.529 min
Delta R.T.: 0.000 min
Response: 25295713
Conc: 714.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.675 min
Delta R.T.: 0.000 min
Response: 40253119
Conc: 710.48 ng/ml




#22 AR-1248-2

R.T.: 5.801 min
Delta R.T.: 0.000 min
Response: 35247884
Conc: 706.45 ng/ml



#22 AR-1248-2


R.T.: 4.909 min
Delta R.T.: 0.000 min
Response: 52563187
Conc: 705.77 ng/ml



#23 AR-1248-3


R.T.: 6.008 min
Delta R.T.: 0.000 min
Response: 39951060
Conc: 708.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.950 min
Delta R.T.: 0.000 min
Response: 54082985
Conc: 703.23 ng/ml




#24 AR-1248-4

R.T.: 6.414 min
Delta R.T.: 0.000 min
Response: 45041663
Conc: 709.93 ng/ml



#24 AR-1248-4


R.T.: 5.121 min
Delta R.T.: 0.000 min
Response: 64278321
Conc: 706.32 ng/ml



#25 AR-1248-5

R.T.: 6.452 min
Delta R.T.: 0.000 min
Response: 43089621
Conc: 713.15 ng/ml

Instrument :
ECD_P
ClientSampleId :



#25 AR-1248-5

R.T.: 5.511 min
Delta R.T.: 0.000 min
Response: 59252859
Conc: 708.99 ng/ml