

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071422\
 Data File : PP049431.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2022 19:37
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 04:58:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 04:58:25 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.354	3.602	71340254	118.8E6	50.000	50.000
2) SA Decachlor...	10.105	8.572	69919216	55954161	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.566	3.814	8560080	15092206	500.000	500.000
9) L2 AR-1221-2	4.654	3.898	6548633	11420945	500.000	500.000
10) L2 AR-1221-3	4.731	3.973	20007763	34023060	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071422\
Data File : PP049431.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2022 19:37
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 04:58:42 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 04:58:25 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.354 min
Delta R.T.: 0.000 min
Response: 71340254
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.602 min
Delta R.T.: 0.000 min
Response: 118757105
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.105 min
Delta R.T.: 0.000 min
Response: 69919216
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.572 min
Delta R.T.: 0.000 min
Response: 55954161
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 4.566 min
Delta R.T.: 0.000 min
Response: 8560080
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 3.814 min
Delta R.T.: 0.000 min
Response: 15092206
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 4.654 min
Delta R.T.: 0.000 min
Response: 6548633
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 3.898 min
Delta R.T.: 0.000 min
Response: 11420945
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.731 min
Delta R.T.: 0.000 min
Response: 20007763
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 3.973 min
Delta R.T.: 0.000 min
Response: 34023060
Conc: 500.00 ng/ml