

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071423\
 Data File : PP058744.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jul 2023 21:45
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 16 05:58:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Jul 02 11:11:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.438	3.669	83116558	128.8E6	51.106	57.352
2)	SA Decachlor...	10.244	8.722	62432536	100.6E6	53.033	55.470
Target Compounds							
16)	L4 AR-1242-1	5.614	4.766	19729416	30518777	497.370	525.476
17)	L4 AR-1242-2	5.637	4.785	28818620	43985544	500.388	534.055
18)	L4 AR-1242-3	5.699	4.962	18339643	23246870	500.902	534.835
19)	L4 AR-1242-4	5.798	5.047	14575994	24373214	500.009	530.051
20)	L4 AR-1242-5	6.539	5.573	14233100	30101634	503.867	526.001

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071423\
Data File : PP058744.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2023 21:45
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 05:58:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Jul 02 11:11:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

