

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071425\
 Data File : PP073777.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jul 2025 16:47
 Operator : YP\AJ
 Sample : Q2578-10
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 WC-A2-10-C

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/15/2025
 Supervised By :mohammad ahmed 07/16/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 02:20:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 08 08:35:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.486	3.778	18928073	28669193	13.821m	15.562m
2) SA Decachlor...	10.174	8.779	10308551	14954048	9.448	11.302

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071425\
Data File : PP073777.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2025 16:47
Operator : YP\AJ
Sample : Q2578-10
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
WC-A2-10-C


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/15/2025
Supervised By :mohammad ahmed 07/16/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 15 02:20:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 08 08:35:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.486 min
Delta R.T.: -0.002 min
Response: 18928073
Conc: 13.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
WC-A2-10-C


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/15/2025
Supervised By :mohammad ahmed 07/16/2025




#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: 0.000 min
Response: 28669193
Conc: 15.56 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.174 min
Delta R.T.: -0.002 min
Response: 10308551
Conc: 9.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.779 min
Delta R.T.: -0.002 min
Response: 14954048
Conc: 11.30 ng/ml