

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071524\
 Data File : PP066338.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2024 22:40
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/16/2024
 Supervised By :Ankita Jodhani 07/16/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 16 08:44:14 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 16 08:20:11 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.792	4.029	8174863	9355074	4.611	4.796
2)	SA Decachlor...	10.795	9.210	8745615	6380462	4.768	4.892
Target Compounds							
26)	L6 AR-1254-1	6.835	5.962	5052926	4730854	54.030	49.710
27)	L6 AR-1254-2	7.052	6.109	7615932	4093643	53.729m	49.291
28)	L6 AR-1254-3	7.422	6.521	6484356	5869805	49.996	46.979
29)	L6 AR-1254-4	7.707	6.750	4396204	3196185	47.942	47.019m
30)	L6 AR-1254-5	8.127	7.173	4569101	4465467	46.099m	44.477

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071524\
Data File : PP066338.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jul 2024 22:40
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/16/2024
Supervised By :Ankita Jodhani 07/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 08:44:14 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 16 08:20:11 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.792 min
Delta R.T.: 0.000 min
Response: 8174863
Conc: 4.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/16/2024
Supervised By :Ankita Jodhani 07/16/2024




#1 Tetrachloro-m-xylene

R.T.: 4.029 min
Delta R.T.: 0.000 min
Response: 9355074
Conc: 4.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.795 min
Delta R.T.: -0.002 min
Response: 8745615
Conc: 4.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.210 min
Delta R.T.: 0.000 min
Response: 6380462
Conc: 4.89 ng/ml




#26 AR-1254-1

R.T.: 6.835 min
Delta R.T.: 0.000 min
Response: 5052926
Conc: 54.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/16/2024
Supervised By :Ankita Jodhani 07/16/2024




#26 AR-1254-1

R.T.: 5.962 min
Delta R.T.: 0.000 min
Response: 4730854
Conc: 49.71 ng/ml




#27 AR-1254-2

R.T.: 7.052 min
Delta R.T.: -0.001 min
Response: 7615932
Conc: 53.73 ng/ml m



#27 AR-1254-2

R.T.: 6.109 min
Delta R.T.: -0.001 min
Response: 4093643
Conc: 49.29 ng/ml




#28 AR-1254-3

R.T.: 7.422 min
Delta R.T.: 0.000 min
Response: 6484356
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/16/2024
Supervised By :Ankita Jodhani 07/16/2024




#28 AR-1254-3

R.T.: 6.521 min
Delta R.T.: -0.001 min
Response: 5869805
Conc: 46.98 ng/ml




#29 AR-1254-4

R.T.: 7.707 min
Delta R.T.: 0.000 min
Response: 4396204
Conc: 47.94 ng/ml



#29 AR-1254-4

R.T.: 6.750 min
Delta R.T.: -0.001 min
Response: 3196185
Conc: 47.02 ng/ml m




#30 AR-1254-5

R.T.: 8.127 min
Delta R.T.: -0.001 min
Response: 4569101
Conc: 46.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/16/2024
Supervised By :Ankita Jodhani 07/16/2024



#30 AR-1254-5

R.T.: 7.173 min
Delta R.T.: -0.001 min
Response: 4465467
Conc: 44.48 ng/ml