

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071525\
 Data File : PP073791.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2025 09:49
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 16 01:41:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 08 08:35:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.483	3.776	79175819	89032052	57.813	48.327
2) SA Decachlor...	10.167	8.773	63941644	70518736	58.605	53.296
Target Compounds						
21) L5 AR-1248-1	5.634	4.853	16586894	22192255	537.305	494.316
22) L5 AR-1248-2	5.905	5.088	22174499	29983740	555.427	487.251
23) L5 AR-1248-3	6.107	5.130	25102106	31127111	565.240	485.871
24) L5 AR-1248-4	6.506	5.302	31033552	36733406	563.422	487.326
25) L5 AR-1248-5	6.545	5.692	29922073	37178628	556.269	498.434



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071525\
Data File : PP073791.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jul 2025 09:49
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 01:41:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 08 08:35:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.483 min
Delta R.T.: -0.005 min
Response: 79175819
Conc: 57.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.776 min
Delta R.T.: -0.003 min
Response: 89032052
Conc: 48.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.167 min
Delta R.T.: -0.009 min
Response: 63941644
Conc: 58.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.773 min
Delta R.T.: -0.008 min
Response: 70518736
Conc: 53.30 ng/ml



#21 AR-1248-1


R.T.: 5.634 min
Delta R.T.: -0.003 min
Response: 16586894
Conc: 537.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.853 min
Delta R.T.: -0.005 min
Response: 22192255
Conc: 494.32 ng/ml




#22 AR-1248-2

R.T.: 5.905 min
Delta R.T.: -0.003 min
Response: 22174499
Conc: 555.43 ng/ml



#22 AR-1248-2


R.T.: 5.088 min
Delta R.T.: -0.006 min
Response: 29983740
Conc: 487.25 ng/ml



#23 AR-1248-3


R.T.: 6.107 min
Delta R.T.: -0.003 min
Response: 25102106
Conc: 565.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.130 min
Delta R.T.: -0.006 min
Response: 31127111
Conc: 485.87 ng/ml




#24 AR-1248-4

R.T.: 6.506 min
Delta R.T.: -0.003 min
Response: 31033552
Conc: 563.42 ng/ml



#24 AR-1248-4


R.T.: 5.302 min
Delta R.T.: -0.006 min
Response: 36733406
Conc: 487.33 ng/ml



#25 AR-1248-5

R.T.: 6.545 min
Delta R.T.: -0.002 min
Response: 29922073
Conc: 556.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.692 min
Delta R.T.: -0.006 min
Response: 37178628
Conc: 498.43 ng/ml