

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071622\
 Data File : PP049519.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2022 16:42
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 17:23:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.351	3.601	75664149	133.8E6	52.141	54.729
2) SA Decachlor...	10.100	8.566	75888661	61354789	53.654	52.683
Target Compounds						
26) L6 AR-1254-1	6.384	5.469	31084314	59722927	493.549	470.886
27) L6 AR-1254-2	6.602	5.615	47799616	51688418	490.643	467.637
28) L6 AR-1254-3	6.973	6.018	50610253	75508099	489.825	463.882
29) L6 AR-1254-4	7.260	6.245	39606255	45226861	494.744	467.427
30) L6 AR-1254-5	7.679	6.661	43410935	62727117	488.410	473.357



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071622\
Data File : PP049519.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jul 2022 16:42
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 15 17:23:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.351 min
Delta R.T.: -0.002 min
Response: 75664149
Conc: 52.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.601 min
Delta R.T.: -0.001 min
Response: 133775004
Conc: 54.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.100 min
Delta R.T.: -0.007 min
Response: 75888661
Conc: 53.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.566 min
Delta R.T.: -0.006 min
Response: 61354789
Conc: 52.68 ng/ml



#26 AR-1254-1


R.T.: 6.384 min
Delta R.T.: -0.002 min
Response: 31084314
Conc: 493.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.469 min
Delta R.T.: -0.002 min
Response: 59722927
Conc: 470.89 ng/ml




#27 AR-1254-2

R.T.: 6.602 min
Delta R.T.: -0.002 min
Response: 47799616
Conc: 490.64 ng/ml



#27 AR-1254-2


R.T.: 5.615 min
Delta R.T.: -0.002 min
Response: 51688418
Conc: 467.64 ng/ml



#28 AR-1254-3


R.T.: 6.973 min
Delta R.T.: -0.001 min
Response: 50610253
Conc: 489.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.018 min
Delta R.T.: -0.002 min
Response: 75508099
Conc: 463.88 ng/ml




#29 AR-1254-4

R.T.: 7.260 min
Delta R.T.: -0.001 min
Response: 39606255
Conc: 494.74 ng/ml



#29 AR-1254-4


R.T.: 6.245 min
Delta R.T.: -0.002 min
Response: 45226861
Conc: 467.43 ng/ml



#30 AR-1254-5

R.T.: 7.679 min
Delta R.T.: -0.002 min
Response: 43410935
Conc: 488.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.661 min
Delta R.T.: -0.002 min
Response: 62727117
Conc: 473.36 ng/ml