

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071725\
 Data File : PP073911.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jul 2025 18:14
 Operator : YP\AJ
 Sample : Q2594-01
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 CC-071325-RW

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/29/2025
 Supervised By :mohammad ahmed 07/29/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 18 01:24:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 08 08:35:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.485	3.780	27112758	72389663	19.797m	39.293m#
2) SA Decachlor...	10.171	8.775	15448391	20161325	14.159	15.237

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071725\
Data File : PP073911.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jul 2025 18:14
Operator : YP\AJ
Sample : Q2594-01
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
CC-071325-RW


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/29/2025
Supervised By :mohammad ahmed 07/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 18 01:24:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 08 08:35:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.485 min
Delta R.T.: -0.003 min
Response: 27112758
Conc: 19.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
CC-071325-RW


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/29/2025
Supervised By :mohammad ahmed 07/29/2025




#1 Tetrachloro-m-xylene

R.T.: 3.780 min
Delta R.T.: 0.001 min
Response: 72389663
Conc: 39.29 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.171 min
Delta R.T.: -0.005 min
Response: 15448391
Conc: 14.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.775 min
Delta R.T.: -0.005 min
Response: 20161325
Conc: 15.24 ng/ml