

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071921\
 Data File : PP037381.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Jul 2021 19:19
 Operator : DD\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 05:59:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.959	3.948	2341531	1414038	51.519	51.675
2) SA Decachlor...	10.991	9.242	2178343	1256497	47.319	45.141
Target Compounds						
21) L5 AR-1248-1	6.279	5.199	505532	288709	481.654	457.493
22) L5 AR-1248-2	6.573	5.463	671954	395928	452.631	482.681
23) L5 AR-1248-3	6.791	5.507	807682	415566	488.970	485.603
24) L5 AR-1248-4	7.220	5.690	910970	497875	481.967	489.228
25) L5 AR-1248-5	7.259	6.113	901920	508048	481.897	478.954



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071921\
Data File : PP037381.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Jul 2021 19:19
Operator : DD\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 05:59:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.959 min
Delta R.T.: -0.001 min
Response: 2341531
Conc: 51.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.948 min
Delta R.T.: -0.002 min
Response: 1414038
Conc: 51.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.991 min
Delta R.T.: -0.001 min
Response: 2178343
Conc: 47.32 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.242 min
Delta R.T.: -0.005 min
Response: 1256497
Conc: 45.14 ng/ml



#21 AR-1248-1


R.T.: 6.279 min
Delta R.T.: -0.001 min
Response: 505532
Conc: 481.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.199 min
Delta R.T.: -0.003 min
Response: 288709
Conc: 457.49 ng/ml




#22 AR-1248-2

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 671954
Conc: 452.63 ng/ml



#22 AR-1248-2


R.T.: 5.463 min
Delta R.T.: -0.002 min
Response: 395928
Conc: 482.68 ng/ml



#23 AR-1248-3


R.T.: 6.791 min
Delta R.T.: 0.000 min
Response: 807682
Conc: 488.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.507 min
Delta R.T.: -0.002 min
Response: 415566
Conc: 485.60 ng/ml




#24 AR-1248-4

R.T.: 7.220 min
Delta R.T.: 0.000 min
Response: 910970
Conc: 481.97 ng/ml



#24 AR-1248-4


R.T.: 5.690 min
Delta R.T.: -0.002 min
Response: 497875
Conc: 489.23 ng/ml



#25 AR-1248-5

R.T.: 7.259 min
Delta R.T.: 0.000 min
Response: 901920
Conc: 481.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.113 min
Delta R.T.: -0.002 min
Response: 508048
Conc: 478.95 ng/ml