

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071923\
 Data File : PP058760.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jul 2023 20:51
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 03:08:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 20 03:06:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.426	3.668	38371945	54702129	25.167	24.795
2) SA Decachlor...	10.221	8.717	23607458	46343336	25.299	25.248
Target Compounds						
16) L4 AR-1242-1	5.601	4.764	10340109	14723490	265.302	252.674
17) L4 AR-1242-2	5.624	4.783	14406111	20736627	262.948	248.687
18) L4 AR-1242-3	5.687	4.960	9307261	11227332	263.734	251.975
19) L4 AR-1242-4	5.785	5.045	7420980	11943371	260.456	256.369
20) L4 AR-1242-5	6.525	5.572	6987838	13757250	264.469	255.159



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP071923\
Data File : PP058760.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jul 2023 20:51
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 03:08:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071923.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 20 03:06:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.426 min
Delta R.T.: 0.000 min
Response: 38371945
Conc: 25.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.668 min
Delta R.T.: 0.000 min
Response: 54702129
Conc: 24.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.221 min
Delta R.T.: -0.002 min
Response: 23607458
Conc: 25.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.717 min
Delta R.T.: 0.000 min
Response: 46343336
Conc: 25.25 ng/ml



#16 AR-1242-1


R.T.: 5.601 min
Delta R.T.: 0.000 min
Response: 10340109
Conc: 265.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.764 min
Delta R.T.: 0.000 min
Response: 14723490
Conc: 252.67 ng/ml




#17 AR-1242-2

R.T.: 5.624 min
Delta R.T.: 0.000 min
Response: 14406111
Conc: 262.95 ng/ml



#17 AR-1242-2


R.T.: 4.783 min
Delta R.T.: 0.000 min
Response: 20736627
Conc: 248.69 ng/ml



#18 AR-1242-3


R.T.: 5.687 min
Delta R.T.: 0.000 min
Response: 9307261
Conc: 263.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 11227332
Conc: 251.98 ng/ml




#19 AR-1242-4

R.T.: 5.785 min
Delta R.T.: 0.000 min
Response: 7420980
Conc: 260.46 ng/ml



#19 AR-1242-4


R.T.: 5.045 min
Delta R.T.: 0.000 min
Response: 11943371
Conc: 256.37 ng/ml



#20 AR-1242-5

R.T.: 6.525 min
Delta R.T.: 0.000 min
Response: 6987838
Conc: 264.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 13757250
Conc: 255.16 ng/ml