

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072018\
 Data File : PP026819.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jul 2018 16:05
 Operator : EI\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 00:40:29 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071318.M
 Quant Title : 8080.M
 QLast Update : Fri Jul 13 23:16:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	12.043	12.156	112.6E6	180.6E6	504.399	452.005
Target Compounds						
6) T MCPP	12.725	0.000	12391274	0	31.422	N.D. #
15) T Picloram	17.596	0.000	203.4E6	0	53.348	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072018\
Data File : PP026819.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jul 2018 16:05
Operator : EI\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 00:40:29 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071318.M
Quant Title : 8080.M
QLast Update : Fri Jul 13 23:16:25 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 12.043 min
Delta R.T.: -0.003 min
Response: 112558138
Conc: 504.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#4 2,4-DCAA

R.T.: 12.156 min
Delta R.T.: -0.003 min
Response: 180587657
Conc: 452.00 ng/ml




#6 MCP

R.T.: 12.725 min
Delta R.T.: 0.018 min
Response: 12391274
Conc: 31.42 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 12.657 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 17.596 min
Delta R.T.: 0.005 min
Response: 203378721
Conc: 53.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 18.258 min
Response: 0
Conc: N.D.