

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072018\
 Data File : PP026840.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Jul 2018 2:31
 Operator : EI\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 03:27:16 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071318.M
 Quant Title : 8080.M
 QLast Update : Fri Jul 13 23:16:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.043	12.156	117.6E6	189.4E6	527.000	473.952
Target Compounds							
6) T	MCPP	12.726	0.000	13604421	0	34.498	N.D. #
15) T	Picloram	17.597	0.000	212.9E6	0	55.852	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072018\
Data File : PP026840.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Jul 2018 2:31
Operator : EI\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 03:27:16 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071318.M
Quant Title : 8080.M
QLast Update : Fri Jul 13 23:16:25 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 12.043 min
Delta R.T.: -0.003 min
Response: 117601718
Conc: 527.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#4 2,4-DCAA

R.T.: 12.156 min
Delta R.T.: -0.002 min
Response: 189356033
Conc: 473.95 ng/ml




#6 MCP

R.T.: 12.726 min
Delta R.T.: 0.018 min
Response: 13604421
Conc: 34.50 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 12.657 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 17.597 min
Delta R.T.: 0.006 min
Response: 212925889
Conc: 55.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 18.258 min
Response: 0
Conc: N.D.