

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072021\
 Data File : PP037439.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jul 2021 18:24
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 02:11:27 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.958	3.948	2499183	1505416	54.988	55.014
2) SA Decachlor...	10.989	9.241	2201997	1309213	47.833	47.035
Target Compounds						
26) L6 AR-1254-1	7.190	6.069	1331359	735699	741.087	495.563 #
27) L6 AR-1254-2	7.418	6.230	1510951	621634	543.180	478.547
28) L6 AR-1254-3	7.800	6.648	1563905	1032869	507.265	484.000
29) L6 AR-1254-4	8.096	6.889	1097593	663571	483.214	489.095
30) L6 AR-1254-5	8.524	7.319	1210921	922837	483.905	492.275



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072021\
Data File : PP037439.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jul 2021 18:24
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 02:11:27 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.958 min
Delta R.T.: -0.002 min
Response: 2499183
Conc: 54.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.948 min
Delta R.T.: -0.002 min
Response: 1505416
Conc: 55.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.989 min
Delta R.T.: -0.003 min
Response: 2201997
Conc: 47.83 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.241 min
Delta R.T.: -0.007 min
Response: 1309213
Conc: 47.03 ng/ml



#26 AR-1254-1


R.T.: 7.190 min
Delta R.T.: 0.000 min
Response: 1331359
Conc: 741.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.069 min
Delta R.T.: -0.004 min
Response: 735699
Conc: 495.56 ng/ml




#27 AR-1254-2

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 1510951
Conc: 543.18 ng/ml



#27 AR-1254-2


R.T.: 6.230 min
Delta R.T.: -0.004 min
Response: 621634
Conc: 478.55 ng/ml



#28 AR-1254-3


R.T.: 7.800 min
Delta R.T.: 0.000 min
Response: 1563905
Conc: 507.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.648 min
Delta R.T.: -0.004 min
Response: 1032869
Conc: 484.00 ng/ml




#29 AR-1254-4

R.T.: 8.096 min
Delta R.T.: 0.000 min
Response: 1097593
Conc: 483.21 ng/ml



#29 AR-1254-4


R.T.: 6.889 min
Delta R.T.: -0.004 min
Response: 663571
Conc: 489.10 ng/ml



#30 AR-1254-5

R.T.: 8.524 min
Delta R.T.: 0.000 min
Response: 1210921
Conc: 483.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.319 min
Delta R.T.: -0.004 min
Response: 922837
Conc: 492.27 ng/ml