

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072021\
 Data File : PP037463.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Jul 2021 1:28
 Operator : DD\AJ
 Sample : M3054-10
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HIA869E-END-1-4

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 04:13:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.961	3.951	591411	358752	13.013	13.110
2)	SA Decachlor...	11.006	9.261	693596	317311	15.067	11.400
Target Compounds							
26)	L6 AR-1254-1	7.191	6.071	1893878	1431537	1054.207	964.276
27)	L6 AR-1254-2	7.421	6.232	2890789	1095859	1039.226	843.615
28)	L6 AR-1254-3	7.802	6.651	2644823	1529244	857.868	716.599
29)	L6 AR-1254-4	8.099	6.892	2241903	1104271	986.996	813.920
30)	L6 AR-1254-5	8.528	7.322	2084420	1239544	832.970	661.218



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072021\
Data File : PP037463.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Jul 2021 1:28
Operator : DD\AJ
Sample : M3054-10
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HIA869E-END-1-4

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 04:13:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.961 min
Delta R.T.: 0.001 min
Response: 591411
Conc: 13.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
HIA869E-END-1-4




#1 Tetrachloro-m-xylene

R.T.: 3.951 min
Delta R.T.: 0.001 min
Response: 358752
Conc: 13.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.006 min
Delta R.T.: 0.013 min
Response: 693596
Conc: 15.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.261 min
Delta R.T.: 0.014 min
Response: 317311
Conc: 11.40 ng/ml



#26 AR-1254-1


R.T.: 7.191 min
Delta R.T.: 0.001 min
Response: 1893878
Conc: 1054.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
HIA869E-END-1-4




#26 AR-1254-1

R.T.: 6.071 min
Delta R.T.: -0.002 min
Response: 1431537
Conc: 964.28 ng/ml




#27 AR-1254-2

R.T.: 7.421 min
Delta R.T.: 0.002 min
Response: 2890789
Conc: 1039.23 ng/ml



#27 AR-1254-2


R.T.: 6.232 min
Delta R.T.: -0.001 min
Response: 1095859
Conc: 843.61 ng/ml



#28 AR-1254-3


R.T.: 7.802 min
Delta R.T.: 0.002 min
Response: 2644823
Conc: 857.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
HIA869E-END-1-4




#28 AR-1254-3

R.T.: 6.651 min
Delta R.T.: -0.002 min
Response: 1529244
Conc: 716.60 ng/ml




#29 AR-1254-4

R.T.: 8.099 min
Delta R.T.: 0.002 min
Response: 2241903
Conc: 987.00 ng/ml



#29 AR-1254-4


R.T.: 6.892 min
Delta R.T.: -0.001 min
Response: 1104271
Conc: 813.92 ng/ml



#30 AR-1254-5

R.T.: 8.528 min
Delta R.T.: 0.002 min
Response: 2084420
Conc: 832.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
HIA869E-END-1-4



#30 AR-1254-5

R.T.: 7.322 min
Delta R.T.: 0.000 min
Response: 1239544
Conc: 661.22 ng/ml