

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072022\
 Data File : PP049680.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jul 2022 07:47
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/21/2022
 Supervised By :Ankita Jodhani 07/21/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 09:14:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.598	75729523	130.1E6	52.186	53.217
2) SA Decachlor...	10.099	8.565	69728096	55077160	49.298	47.293
Target Compounds						
3) L1 AR-1016-1	5.526	4.671	28957876	43136425	512.555	483.919
4) L1 AR-1016-2	5.549	4.689	41609051	61140055	517.479	486.009
5) L1 AR-1016-3	5.611	4.864	25243451	31966728	515.455	486.451
6) L1 AR-1016-4	5.711	4.905	21649885	25081037	535.003	471.814
7) L1 AR-1016-5	6.005	5.117	20496350	34107954	500.770	496.067
31) L7 AR-1260-1	7.134	6.145	39278360	46548460	519.474	444.162
32) L7 AR-1260-2	7.392	6.331	44000682	54853529	464.806m	455.875
33) L7 AR-1260-3	7.750	6.484	28239888	49595685	450.969m	438.303
34) L7 AR-1260-4	7.978	6.955	36287002	33991480	468.977m	444.380
35) L7 AR-1260-5	8.298	7.196	74710249	76998919	485.920m	450.289

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072022\
Data File : PP049680.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jul 2022 07:47
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/21/2022
Supervised By :Ankita Jodhani 07/21/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 09:14:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

