

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072022\
 Data File : PP049629.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2022 14:15
 Operator : YP\AJ
 Sample : N3776-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 72-11966

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/21/2022
 Supervised By :Ankita Jodhani 07/21/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 03:35:17 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.352	3.601	30797140	52243875	21.223m	21.373
2) SA Decachlor...	10.105	8.570	32498926	23698426	22.977m	20.349

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072022\
Data File : PP049629.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2022 14:15
Operator : YP\AJ
Sample : N3776-01
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
72-11966


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/21/2022
Supervised By :Ankita Jodhani 07/21/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 03:35:17 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.352 min
Delta R.T.: -0.001 min
Response: 30797140
Conc: 21.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
72-11966


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/21/2022
Supervised By :Ankita Jodhani 07/21/2022




#1 Tetrachloro-m-xylene

R.T.: 3.601 min
Delta R.T.: 0.000 min
Response: 52243875
Conc: 21.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.105 min
Delta R.T.: -0.002 min
Response: 32498926
Conc: 22.98 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.570 min
Delta R.T.: -0.003 min
Response: 23698426
Conc: 20.35 ng/ml