

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072121\
 Data File : PP037482.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Jul 2021 14:49
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 15:35:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.958	3.948	2443152	1423167	53.755	52.008
2) SA Decachlor...	10.991	9.242	2148092	1154925	46.662	41.492
Target Compounds						
3) L1 AR-1016-1	6.279	5.198	891877	507260	513.248	490.865
4) L1 AR-1016-2	6.303	5.218	1264681	705848	505.854	492.996
5) L1 AR-1016-3	6.368	5.409	801315	389841	503.120	498.963
6) L1 AR-1016-4	6.476	5.462	671431	298327	509.205	494.254
7) L1 AR-1016-5	6.790	5.689	655882	373103	518.481	483.172
31) L7 AR-1260-1	7.966	6.787	1043705	646707	474.102	468.014
32) L7 AR-1260-2	8.232	6.985	1262059	766209	453.689	456.094
33) L7 AR-1260-3	8.597	7.139	993478	733807	491.534	428.274
34) L7 AR-1260-4	8.828	7.622	1182706	621762	480.811	457.336
35) L7 AR-1260-5	9.157	7.872	2195631	1437273	475.822	445.607



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072121\
Data File : PP037482.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Jul 2021 14:49
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 15:35:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.958 min
Delta R.T.: -0.002 min
Response: 2443152
Conc: 53.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.948 min
Delta R.T.: -0.002 min
Response: 1423167
Conc: 52.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.991 min
Delta R.T.: -0.002 min
Response: 2148092
Conc: 46.66 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.242 min
Delta R.T.: -0.006 min
Response: 1154925
Conc: 41.49 ng/ml



#3 AR-1016-1


R.T.: 6.279 min
Delta R.T.: -0.002 min
Response: 891877
Conc: 513.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.198 min
Delta R.T.: -0.003 min
Response: 507260
Conc: 490.86 ng/ml




#4 AR-1016-2

R.T.: 6.303 min
Delta R.T.: 0.000 min
Response: 1264681
Conc: 505.85 ng/ml



#4 AR-1016-2


R.T.: 5.218 min
Delta R.T.: -0.004 min
Response: 705848
Conc: 493.00 ng/ml



#5 AR-1016-3


R.T.: 6.368 min
Delta R.T.: -0.002 min
Response: 801315
Conc: 503.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.409 min
Delta R.T.: -0.004 min
Response: 389841
Conc: 498.96 ng/ml




#6 AR-1016-4

R.T.: 6.476 min
Delta R.T.: -0.001 min
Response: 671431
Conc: 509.20 ng/ml



#6 AR-1016-4


R.T.: 5.462 min
Delta R.T.: -0.004 min
Response: 298327
Conc: 494.25 ng/ml



#7 AR-1016-5


R.T.: 6.790 min
Delta R.T.: -0.002 min
Response: 655882
Conc: 518.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.689 min
Delta R.T.: -0.003 min
Response: 373103
Conc: 483.17 ng/ml




#31 AR-1260-1

R.T.: 7.966 min
Delta R.T.: -0.002 min
Response: 1043705
Conc: 474.10 ng/ml



#31 AR-1260-1


R.T.: 6.787 min
Delta R.T.: -0.004 min
Response: 646707
Conc: 468.01 ng/ml



#32 AR-1260-2


R.T.: 8.232 min
Delta R.T.: -0.002 min
Response: 1262059
Conc: 453.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.985 min
Delta R.T.: -0.004 min
Response: 766209
Conc: 456.09 ng/ml




#33 AR-1260-3

R.T.: 8.597 min
Delta R.T.: -0.001 min
Response: 993478
Conc: 491.53 ng/ml



#33 AR-1260-3


R.T.: 7.139 min
Delta R.T.: -0.004 min
Response: 733807
Conc: 428.27 ng/ml



#34 AR-1260-4


R.T.: 8.828 min
Delta R.T.: -0.001 min
Response: 1182706
Conc: 480.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.622 min
Delta R.T.: -0.004 min
Response: 621762
Conc: 457.34 ng/ml



#35 AR-1260-5

R.T.: 9.157 min
Delta R.T.: -0.001 min
Response: 2195631
Conc: 475.82 ng/ml



#35 AR-1260-5

R.T.: 7.872 min
Delta R.T.: -0.004 min
Response: 1437273
Conc: 445.61 ng/ml