

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072122\
 Data File : PP049702.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jul 2022 15:08
 Operator : YP\AJ
 Sample : N3810-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 COMP-3

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 16:35:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.353	3.602	31369379	51462966	21.617	21.054
2) SA Decachlor...	10.106	8.570	20604581	16251853	14.568	13.955

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072122\
Data File : PP049702.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jul 2022 15:08
Operator : YP\AJ
Sample : N3810-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
COMP-3

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 16:35:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.353 min
Delta R.T.: 0.000 min
Response: 31369379
Conc: 21.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-3




#1 Tetrachloro-m-xylene

R.T.: 3.602 min
Delta R.T.: 0.000 min
Response: 51462966
Conc: 21.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.106 min
Delta R.T.: -0.001 min
Response: 20604581
Conc: 14.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.570 min
Delta R.T.: -0.003 min
Response: 16251853
Conc: 13.95 ng/ml