

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072123\
 Data File : PP058845.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jul 2023 16:47
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 17:25:33 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 21 17:23:34 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.425	3.664	6553557	8746805	4.625	4.384
2) SA Decachlor...	10.227	8.715	4068252	7149359	4.446	4.781
Target Compounds						
16) L4 AR-1242-1	5.603	4.762	1834745	2450823	47.454	48.701
17) L4 AR-1242-2	5.626	4.781	2600084	3388412	47.244	46.291
18) L4 AR-1242-3	5.688	4.959	1642842	1666046	45.770	42.305
19) L4 AR-1242-4	5.786	5.043	1318094	1856576	45.075	44.661
20) L4 AR-1242-5	6.528	5.570	1352528	2471513	45.645	51.192



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072123\
Data File : PP058845.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jul 2023 16:47
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 17:25:33 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 21 17:23:34 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.425 min
Delta R.T.: 0.001 min
Response: 6553557
Conc: 4.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.664 min
Delta R.T.: 0.000 min
Response: 8746805
Conc: 4.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.227 min
Delta R.T.: 0.003 min
Response: 4068252
Conc: 4.45 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.715 min
Delta R.T.: 0.000 min
Response: 7149359
Conc: 4.78 ng/ml



#16 AR-1242-1


R.T.: 5.603 min
Delta R.T.: 0.002 min
Response: 1834745
Conc: 47.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#16 AR-1242-1

R.T.: 4.762 min
Delta R.T.: 0.000 min
Response: 2450823
Conc: 48.70 ng/ml




#17 AR-1242-2

R.T.: 5.626 min
Delta R.T.: 0.001 min
Response: 2600084
Conc: 47.24 ng/ml



#17 AR-1242-2


R.T.: 4.781 min
Delta R.T.: 0.000 min
Response: 3388412
Conc: 46.29 ng/ml



#18 AR-1242-3


R.T.: 5.688 min
Delta R.T.: 0.002 min
Response: 1642842
Conc: 45.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#18 AR-1242-3

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 1666046
Conc: 42.30 ng/ml




#19 AR-1242-4

R.T.: 5.786 min
Delta R.T.: 0.000 min
Response: 1318094
Conc: 45.07 ng/ml



#19 AR-1242-4


R.T.: 5.043 min
Delta R.T.: 0.000 min
Response: 1856576
Conc: 44.66 ng/ml



#20 AR-1242-5

R.T.: 6.528 min
Delta R.T.: 0.002 min
Response: 1352528
Conc: 45.64 ng/ml

Instrument :
ECD_P
ClientSampleId :



#20 AR-1242-5

R.T.: 5.570 min
Delta R.T.: 0.000 min
Response: 2471513
Conc: 51.19 ng/ml