

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072125\
 Data File : PP074016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jul 2025 00:17
 Operator : YP\AJ
 Sample : Q2663-14 10X
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HEH700H-1-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 22 02:20:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 08 08:35:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.479	3.772	1754467	2925753	1.281	1.588
2)	SA Decachlor...	10.166	8.769	1782359	2134827	1.634	1.613
Target Compounds							
3)	L1 AR-1016-1	5.637	4.848	136178	31291	2.866	0.459 #
4)	L1 AR-1016-2	5.651	4.848	152531	31291	2.142	0.307 #
5)	L1 AR-1016-3	5.716	5.050	75249	27880	1.723	0.515 #
6)	L1 AR-1016-4	5.830	5.085	48910	1531375	1.361	34.898 #
7)	L1 AR-1016-5	6.104	5.298	331300	1336877	10.574	24.493 #
8)	L2 AR-1221-1	0.000	3.985	0	29274	N.D.	1.082 #
9)	L2 AR-1221-2	4.816f	4.095	27552	28644	2.040	1.407 #
10)	L2 AR-1221-3	4.860	4.174	71760	27242	1.725	0.443 #
11)	L3 AR-1232-1	4.860	4.174	71760	27242	2.198	0.584 #
12)	L3 AR-1232-2	5.379	4.848	55317	31291	3.403	0.657 #
13)	L3 AR-1232-3	5.651	5.050	152531	27880	4.623	1.114 #
14)	L3 AR-1232-4	5.830	5.126	48910	518215	2.974	23.929 #
15)	L3 AR-1232-5	5.902	5.298	818502	1336877	77.434	59.250
16)	L4 AR-1242-1	5.637	4.848	136178	31291	3.566	0.542 #
17)	L4 AR-1242-2	5.651	4.848	152531	31291	2.547	0.362 #
18)	L4 AR-1242-3	5.716	5.050	75249	27880	2.055	0.607 #
19)	L4 AR-1242-4	5.830	5.126	48910	518215	1.522	11.708 #
20)	L4 AR-1242-5	6.544	5.649	318896	3810110	9.741	68.971 #
21)	L5 AR-1248-1	5.637	4.848	136178	31291	4.411	0.697 #
22)	L5 AR-1248-2	5.902	5.085	818502	1531375	20.502	24.886
23)	L5 AR-1248-3	6.104	5.126	331300	518215	7.460	8.089
24)	L5 AR-1248-4	6.500	5.298	546161	1336877	9.916	17.736 #
25)	L5 AR-1248-5	6.544	5.689	318896	681141	5.928	9.132 #
26)	L6 AR-1254-1	6.478	5.649	1203936	3810110	22.466	32.837 #
27)	L6 AR-1254-2	6.695	5.796	2247413	3268257	26.983	32.480
28)	L6 AR-1254-3	7.057	6.197	2306228	4155828	26.109	26.908
29)	L6 AR-1254-4	7.340	6.426	1936816	3029253	24.620	32.036 #
30)	L6 AR-1254-5	7.757	6.841	2115214	4046628	28.708	30.324
31)	L7 AR-1260-1	7.223	6.329	1096832	2117660	18.603	21.706
32)	L7 AR-1260-2	7.476	6.516	1578338	2337029	16.468	19.025
33)	L7 AR-1260-3	7.831	6.666	604253	1797645	8.269	16.422 #
34)	L7 AR-1260-4	8.047	7.136	1276622	701980	19.550	7.851 #
35)	L7 AR-1260-5	8.375	7.379	919796	1148895	6.097	5.112
36)	L8 AR-1262-1	8.047	6.912	1276622	345869	14.669	2.342 #
37)	L8 AR-1262-2	8.375	7.136	919796	701980	4.645	5.492
38)	L8 AR-1262-3	8.695	7.660	368420	461567	2.930	4.048 #
39)	L8 AR-1262-4	8.783	7.722	80625	814965	0.864	4.428 #
40)	L8 AR-1262-5	9.414	8.223	109063	176087	1.721	2.085
41)	L9 AR-1268-1	8.695	7.660	368420	461567	1.650	1.502

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072125\
Data File : PP074016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jul 2025 00:17
Operator : YP\AJ
Sample : Q2663-14 10X
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 22 02:20:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 08 08:35:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.783	7.722	80625	814965	0.423	3.067 #
43)	L9 AR-1268-3	8.998	7.947	172949	367158	1.045	1.628 #
44)	L9 AR-1268-4	9.414	8.223	109063	176087	1.568	1.894
45)	L9 AR-1268-5	9.832	8.544	-47488	85519	N.D.	0.139 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072125\
Data File : PP074016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jul 2025 00:17
Operator : YP\AJ
Sample : Q2663-14 10X
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 22 02:20:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 08 08:35:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.479 min
Delta R.T.: -0.009 min
Response: 1754467
Conc: 1.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#1 Tetrachloro-m-xylene

R.T.: 3.772 min
Delta R.T.: -0.006 min
Response: 2925753
Conc: 1.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.166 min
Delta R.T.: -0.010 min
Response: 1782359
Conc: 1.63 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.769 min
Delta R.T.: -0.011 min
Response: 2134827
Conc: 1.61 ng/ml



#3 AR-1016-1


R.T.: 5.637 min
Delta R.T.: -0.002 min
Response: 136178
Conc: 2.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#3 AR-1016-1

R.T.: 4.848 min
Delta R.T.: -0.010 min
Response: 31291
Conc: 0.46 ng/ml




#4 AR-1016-2

R.T.: 5.651 min
Delta R.T.: -0.009 min
Response: 152531
Conc: 2.14 ng/ml



#4 AR-1016-2


R.T.: 4.848 min
Delta R.T.: -0.027 min
Response: 31291
Conc: 0.31 ng/ml



#5 AR-1016-3


R.T.: 5.716 min
Delta R.T.: -0.007 min
Response: 75249
Conc: 1.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#5 AR-1016-3

R.T.: 5.050 min
Delta R.T.: -0.002 min
Response: 27880
Conc: 0.52 ng/ml




#6 AR-1016-4

R.T.: 5.830 min
Delta R.T.: 0.010 min
Response: 48910
Conc: 1.36 ng/ml



#6 AR-1016-4


R.T.: 5.085 min
Delta R.T.: -0.009 min
Response: 1531375
Conc: 34.90 ng/ml



#7 AR-1016-5


R.T.: 6.104 min
Delta R.T.: -0.008 min
Response: 331300
Conc: 10.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#7 AR-1016-5

R.T.: 5.298 min
Delta R.T.: -0.009 min
Response: 1336877
Conc: 24.49 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.690 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.985 min
Delta R.T.: -0.003 min
Response: 29274
Conc: 1.08 ng/ml



#9 AR-1221-2


R.T.: 4.816 min
Delta R.T.: 0.041 min
Response: 27552
Conc: 2.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#9 AR-1221-2

R.T.: 4.095 min
Delta R.T.: 0.021 min
Response: 28644
Conc: 1.41 ng/ml




#10 AR-1221-3

R.T.: 4.860 min
Delta R.T.: 0.009 min
Response: 71760
Conc: 1.72 ng/ml



#10 AR-1221-3


R.T.: 4.174 min
Delta R.T.: 0.024 min
Response: 27242
Conc: 0.44 ng/ml



#11 AR-1232-1


R.T.: 4.860 min
Delta R.T.: 0.010 min
Response: 71760
Conc: 2.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#11 AR-1232-1

R.T.: 4.174 min
Delta R.T.: 0.024 min
Response: 27242
Conc: 0.58 ng/ml




#12 AR-1232-2

R.T.: 5.379 min
Delta R.T.: 0.005 min
Response: 55317
Conc: 3.40 ng/ml



#12 AR-1232-2


R.T.: 4.848 min
Delta R.T.: -0.027 min
Response: 31291
Conc: 0.66 ng/ml



#13 AR-1232-3


R.T.: 5.651 min
Delta R.T.: -0.009 min
Response: 152531
Conc: 4.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#13 AR-1232-3

R.T.: 5.050 min
Delta R.T.: -0.002 min
Response: 27880
Conc: 1.11 ng/ml




#14 AR-1232-4

R.T.: 5.830 min
Delta R.T.: 0.010 min
Response: 48910
Conc: 2.97 ng/ml



#14 AR-1232-4


R.T.: 5.126 min
Delta R.T.: -0.011 min
Response: 518215
Conc: 23.93 ng/ml



#15 AR-1232-5


R.T.: 5.902 min
Delta R.T.: -0.007 min
Response: 818502
Conc: 77.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#15 AR-1232-5

R.T.: 5.298 min
Delta R.T.: -0.009 min
Response: 1336877
Conc: 59.25 ng/ml




#16 AR-1242-1

R.T.: 5.637 min
Delta R.T.: 0.001 min
Response: 136178
Conc: 3.57 ng/ml



#16 AR-1242-1


R.T.: 4.848 min
Delta R.T.: -0.010 min
Response: 31291
Conc: 0.54 ng/ml



#17 AR-1242-2


R.T.: 5.651 min
Delta R.T.: -0.007 min
Response: 152531
Conc: 2.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#17 AR-1242-2

R.T.: 4.848 min
Delta R.T.: -0.027 min
Response: 31291
Conc: 0.36 ng/ml




#18 AR-1242-3

R.T.: 5.716 min
Delta R.T.: -0.004 min
Response: 75249
Conc: 2.05 ng/ml



#18 AR-1242-3


R.T.: 5.050 min
Delta R.T.: -0.002 min
Response: 27880
Conc: 0.61 ng/ml



#19 AR-1242-4


R.T.: 5.830 min
Delta R.T.: 0.013 min
Response: 48910
Conc: 1.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#19 AR-1242-4

R.T.: 5.126 min
Delta R.T.: -0.010 min
Response: 518215
Conc: 11.71 ng/ml




#20 AR-1242-5

R.T.: 6.544 min
Delta R.T.: -0.002 min
Response: 318896
Conc: 9.74 ng/ml



#20 AR-1242-5


R.T.: 5.649 min
Delta R.T.: -0.008 min
Response: 3810110
Conc: 68.97 ng/ml



#21 AR-1248-1


R.T.: 5.637 min
Delta R.T.: 0.000 min
Response: 136178
Conc: 4.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#21 AR-1248-1

R.T.: 4.848 min
Delta R.T.: -0.009 min
Response: 31291
Conc: 0.70 ng/ml




#22 AR-1248-2

R.T.: 5.902 min
Delta R.T.: -0.005 min
Response: 818502
Conc: 20.50 ng/ml



#22 AR-1248-2


R.T.: 5.085 min
Delta R.T.: -0.009 min
Response: 1531375
Conc: 24.89 ng/ml



#23 AR-1248-3


R.T.: 6.104 min
Delta R.T.: -0.006 min
Response: 331300
Conc: 7.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#23 AR-1248-3

R.T.: 5.126 min
Delta R.T.: -0.010 min
Response: 518215
Conc: 8.09 ng/ml




#24 AR-1248-4

R.T.: 6.500 min
Delta R.T.: -0.008 min
Response: 546161
Conc: 9.92 ng/ml



#24 AR-1248-4


R.T.: 5.298 min
Delta R.T.: -0.009 min
Response: 1336877
Conc: 17.74 ng/ml



#25 AR-1248-5


R.T.: 6.544 min
Delta R.T.: -0.003 min
Response: 318896
Conc: 5.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#25 AR-1248-5

R.T.: 5.689 min
Delta R.T.: -0.009 min
Response: 681141
Conc: 9.13 ng/ml




#26 AR-1254-1

R.T.: 6.478 min
Delta R.T.: -0.007 min
Response: 1203936
Conc: 22.47 ng/ml



#26 AR-1254-1


R.T.: 5.649 min
Delta R.T.: -0.009 min
Response: 3810110
Conc: 32.84 ng/ml



#27 AR-1254-2


R.T.: 6.695 min
Delta R.T.: -0.007 min
Response: 2247413
Conc: 26.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#27 AR-1254-2

R.T.: 5.796 min
Delta R.T.: -0.011 min
Response: 3268257
Conc: 32.48 ng/ml




#28 AR-1254-3

R.T.: 7.057 min
Delta R.T.: -0.007 min
Response: 2306228
Conc: 26.11 ng/ml



#28 AR-1254-3


R.T.: 6.197 min
Delta R.T.: -0.011 min
Response: 4155828
Conc: 26.91 ng/ml



#29 AR-1254-4


R.T.: 7.340 min
Delta R.T.: -0.007 min
Response: 1936816
Conc: 24.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#29 AR-1254-4

R.T.: 6.426 min
Delta R.T.: -0.010 min
Response: 3029253
Conc: 32.04 ng/ml




#30 AR-1254-5

R.T.: 7.757 min
Delta R.T.: -0.006 min
Response: 2115214
Conc: 28.71 ng/ml



#30 AR-1254-5


R.T.: 6.841 min
Delta R.T.: -0.011 min
Response: 4046628
Conc: 30.32 ng/ml



#31 AR-1260-1


R.T.: 7.223 min
Delta R.T.: -0.006 min
Response: 1096832
Conc: 18.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#31 AR-1260-1

R.T.: 6.329 min
Delta R.T.: -0.008 min
Response: 2117660
Conc: 21.71 ng/ml




#32 AR-1260-2

R.T.: 7.476 min
Delta R.T.: -0.007 min
Response: 1578338
Conc: 16.47 ng/ml



#32 AR-1260-2


R.T.: 6.516 min
Delta R.T.: -0.009 min
Response: 2337029
Conc: 19.03 ng/ml



#33 AR-1260-3


R.T.: 7.831 min
Delta R.T.: -0.009 min
Response: 604253
Conc: 8.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#33 AR-1260-3

R.T.: 6.666 min
Delta R.T.: -0.011 min
Response: 1797645
Conc: 16.42 ng/ml




#34 AR-1260-4

R.T.: 8.047 min
Delta R.T.: -0.018 min
Response: 1276622
Conc: 19.55 ng/ml



#34 AR-1260-4


R.T.: 7.136 min
Delta R.T.: -0.010 min
Response: 701980
Conc: 7.85 ng/ml



#35 AR-1260-5


R.T.: 8.375 min
Delta R.T.: -0.008 min
Response: 919796
Conc: 6.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#35 AR-1260-5

R.T.: 7.379 min
Delta R.T.: -0.010 min
Response: 1148895
Conc: 5.11 ng/ml




#36 AR-1262-1

R.T.: 8.047 min
Delta R.T.: -0.019 min
Response: 1276622
Conc: 14.67 ng/ml



#36 AR-1262-1


R.T.: 6.912 min
Delta R.T.: 0.023 min
Response: 345869
Conc: 2.34 ng/ml



#37 AR-1262-2


R.T.: 8.375 min
Delta R.T.: -0.008 min
Response: 919796
Conc: 4.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#37 AR-1262-2

R.T.: 7.136 min
Delta R.T.: -0.011 min
Response: 701980
Conc: 5.49 ng/ml




#38 AR-1262-3

R.T.: 8.695 min
Delta R.T.: -0.002 min
Response: 368420
Conc: 2.93 ng/ml



#38 AR-1262-3


R.T.: 7.660 min
Delta R.T.: -0.009 min
Response: 461567
Conc: 4.05 ng/ml



#39 AR-1262-4


R.T.: 8.783 min
Delta R.T.: 0.000 min
Response: 80625
Conc: 0.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#39 AR-1262-4

R.T.: 7.722 min
Delta R.T.: -0.013 min
Response: 814965
Conc: 4.43 ng/ml




#40 AR-1262-5

R.T.: 9.414 min
Delta R.T.: -0.018 min
Response: 109063
Conc: 1.72 ng/ml



#40 AR-1262-5


R.T.: 8.223 min
Delta R.T.: -0.009 min
Response: 176087
Conc: 2.09 ng/ml



#41 AR-1268-1


R.T.: 8.695 min
Delta R.T.: 0.002 min
Response: 368420
Conc: 1.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#41 AR-1268-1

R.T.: 7.660 min
Delta R.T.: -0.009 min
Response: 461567
Conc: 1.50 ng/ml




#42 AR-1268-2

R.T.: 8.783 min
Delta R.T.: -0.003 min
Response: 80625
Conc: 0.42 ng/ml



#42 AR-1268-2


R.T.: 7.722 min
Delta R.T.: -0.013 min
Response: 814965
Conc: 3.07 ng/ml



#43 AR-1268-3


R.T.: 8.998 min
Delta R.T.: -0.016 min
Response: 172949
Conc: 1.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2




#43 AR-1268-3

R.T.: 7.947 min
Delta R.T.: 0.011 min
Response: 367158
Conc: 1.63 ng/ml




#44 AR-1268-4

R.T.: 9.414 min
Delta R.T.: -0.017 min
Response: 109063
Conc: 1.57 ng/ml



#44 AR-1268-4


R.T.: 8.223 min
Delta R.T.: -0.008 min
Response: 176087
Conc: 1.89 ng/ml



#45 AR-1268-5

R.T.: 9.832 min
Delta R.T.: -0.009 min
Response: -47488
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
HEH700H-1-2



#45 AR-1268-5

R.T.: 8.544 min
Delta R.T.: 0.018 min
Response: 85519
Conc: 0.14 ng/ml