

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072222\
 Data File : PP049740.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jul 2022 17:19
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 02:29:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.352	3.602	77397288	138.6E6	53.336	56.704
2)	SA Decachlor...	10.106	8.569	69955496	55513077	49.459	47.667
Target Compounds							
26)	L6 AR-1254-1	6.386	5.469	29963833	56434486	475.758	444.959
27)	L6 AR-1254-2	6.605	5.616	45619019	48101273	468.260	435.183
28)	L6 AR-1254-3	6.976	6.018	48242465	69781406	466.908	428.700
29)	L6 AR-1254-4	7.263	6.246	36660583	42072603	457.948	434.827
30)	L6 AR-1254-5	7.683	6.662	46609900	56868391	524.402	429.145



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072222\
Data File : PP049740.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jul 2022 17:19
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 02:29:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.352 min
Delta R.T.: 0.000 min
Response: 77397288
Conc: 53.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.602 min
Delta R.T.: 0.000 min
Response: 138604625
Conc: 56.70 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.106 min
Delta R.T.: 0.000 min
Response: 69955496
Conc: 49.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.569 min
Delta R.T.: -0.003 min
Response: 55513077
Conc: 47.67 ng/ml



#26 AR-1254-1


R.T.: 6.386 min
Delta R.T.: 0.000 min
Response: 29963833
Conc: 475.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.469 min
Delta R.T.: -0.002 min
Response: 56434486
Conc: 444.96 ng/ml




#27 AR-1254-2

R.T.: 6.605 min
Delta R.T.: 0.000 min
Response: 45619019
Conc: 468.26 ng/ml



#27 AR-1254-2


R.T.: 5.616 min
Delta R.T.: -0.001 min
Response: 48101273
Conc: 435.18 ng/ml



#28 AR-1254-3


R.T.: 6.976 min
Delta R.T.: 0.001 min
Response: 48242465
Conc: 466.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.018 min
Delta R.T.: -0.001 min
Response: 69781406
Conc: 428.70 ng/ml




#29 AR-1254-4

R.T.: 7.263 min
Delta R.T.: 0.001 min
Response: 36660583
Conc: 457.95 ng/ml



#29 AR-1254-4


R.T.: 6.246 min
Delta R.T.: 0.000 min
Response: 42072603
Conc: 434.83 ng/ml



#30 AR-1254-5

R.T.: 7.683 min
Delta R.T.: 0.002 min
Response: 46609900
Conc: 524.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.662 min
Delta R.T.: 0.000 min
Response: 56868391
Conc: 429.14 ng/ml