

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072321\
 Data File : PP037645.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Jul 2021 21:38
 Operator : DD\AJ
 Sample : M3152-10
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 COLL-HOLE-B

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 07:39:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.955	3.946	994331	590176	21.878	21.567
2)	SA Decachlor...	10.981	9.234	879870	404761	19.113	14.542
Target Compounds							
8)	L2 AR-1221-1	5.206	0.000	40618	0	76.583	N.D. #
9)	L2 AR-1221-2	5.307	0.000	11838	0	29.220	N.D. #
28)	L6 AR-1254-3	7.812	0.000	10388	0	3.369	N.D. #
30)	L6 AR-1254-5	8.545f	7.313	11979	11100	4.787	5.921
32)	L7 AR-1260-2	8.227	0.000	8850	0	3.181	N.D. #
33)	L7 AR-1260-3	8.591	0.000	10760	0	5.323	N.D. #
34)	L7 AR-1260-4	8.823	0.000	19822	0	8.058	N.D. #
35)	L7 AR-1260-5	9.151	7.867	25404	12078	5.505	3.745 #
36)	L8 AR-1262-1	8.591	7.313	10760	11100	3.694	10.690 #
37)	L8 AR-1262-2	9.151	7.867	25404	12078	4.837	3.472 #
44)	L9 AR-1268-4	10.266f	0.000	84136	0	36.777	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072321\
Data File : PP037645.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Jul 2021 21:38
Operator : DD\AJ
Sample : M3152-10
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
COLL-HOLE-B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 26 07:39:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.955 min
Delta R.T.: -0.005 min
Response: 994331
Conc: 21.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
COLL-HOLE-B




#1 Tetrachloro-m-xylene

R.T.: 3.946 min
Delta R.T.: -0.004 min
Response: 590176
Conc: 21.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.981 min
Delta R.T.: -0.011 min
Response: 879870
Conc: 19.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.234 min
Delta R.T.: -0.013 min
Response: 404761
Conc: 14.54 ng/ml



#8 AR-1221-1


R.T.: 5.206 min
Delta R.T.: -0.004 min
Response: 40618
Conc: 76.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
COLL-HOLE-B




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.207 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.307 min
Delta R.T.: -0.002 min
Response: 11838
Conc: 29.22 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.305 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.812 min
Delta R.T.: 0.012 min
Response: 10388
Conc: 3.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
COLL-HOLE-B




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.652 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 8.545 min
Delta R.T.: 0.020 min
Response: 11979
Conc: 4.79 ng/ml



#30 AR-1254-5


R.T.: 7.313 min
Delta R.T.: -0.010 min
Response: 11100
Conc: 5.92 ng/ml



#32 AR-1260-2


R.T.: 8.227 min
Delta R.T.: -0.007 min
Response: 8850
Conc: 3.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
COLL-HOLE-B




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 6.989 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 8.591 min
Delta R.T.: -0.008 min
Response: 10760
Conc: 5.32 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 7.143 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 8.823 min
Delta R.T.: -0.007 min
Response: 19822
Conc: 8.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
COLL-HOLE-B




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 7.627 min
Response: 0
Conc: N.D.




#35 AR-1260-5

R.T.: 9.151 min
Delta R.T.: -0.007 min
Response: 25404
Conc: 5.51 ng/ml



#35 AR-1260-5


R.T.: 7.867 min
Delta R.T.: -0.009 min
Response: 12078
Conc: 3.74 ng/ml



#36 AR-1262-1


R.T.: 8.591 min
Delta R.T.: -0.007 min
Response: 10760
Conc: 3.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
COLL-HOLE-B




#36 AR-1262-1

R.T.: 7.313 min
Delta R.T.: -0.008 min
Response: 11100
Conc: 10.69 ng/ml




#37 AR-1262-2

R.T.: 9.151 min
Delta R.T.: -0.006 min
Response: 25404
Conc: 4.84 ng/ml



#37 AR-1262-2


R.T.: 7.867 min
Delta R.T.: -0.008 min
Response: 12078
Conc: 3.47 ng/ml



#44 AR-1268-4

R.T.: 10.266 min
Delta R.T.: 0.030 min
Response: 84136
Conc: 36.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
COLL-HOLE-B



#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 8.725 min
Response: 0
Conc: N.D.