

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072423\
 Data File : PP058897.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2023 13:54
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 21:31:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 22 02:26:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.428	3.667	70226640	106.0E6	51.011	55.143
2) SA Decachlor...	10.228	8.712	48146931	67468251	55.889	47.683
Target Compounds						
3) L1 AR-1016-1	5.603	4.762	23586162	32650178	540.150	568.060
4) L1 AR-1016-2	5.626	4.781	33429481	47178505	536.891	573.625
5) L1 AR-1016-3	5.688	4.958	21550668	24810660	538.254	564.198
6) L1 AR-1016-4	5.787	5.001	17462766	20403149	541.066	524.697
7) L1 AR-1016-5	6.083	5.214	18148958	27282102	532.139	551.527
31) L7 AR-1260-1	7.216	6.254	27845799	49735884	480.140	502.515
32) L7 AR-1260-2	7.473	6.444	30652889	58678978	505.878	516.118
33) L7 AR-1260-3	7.834	6.598	22638557	52659086	511.328	481.134
34) L7 AR-1260-4	8.059	7.072	25280173	35996931	512.557	423.331
35) L7 AR-1260-5	8.382	7.314	47753221	76192407	535.589	449.299



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072423\
Data File : PP058897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2023 13:54
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 21:31:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 02:26:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.428 min
Delta R.T.: 0.003 min
Response: 70226640
Conc: 51.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.667 min
Delta R.T.: 0.003 min
Response: 106046079
Conc: 55.14 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.228 min
Delta R.T.: 0.001 min
Response: 48146931
Conc: 55.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.712 min
Delta R.T.: -0.002 min
Response: 67468251
Conc: 47.68 ng/ml



#3 AR-1016-1


R.T.: 5.603 min
Delta R.T.: 0.001 min
Response: 23586162
Conc: 540.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.762 min
Delta R.T.: 0.000 min
Response: 32650178
Conc: 568.06 ng/ml




#4 AR-1016-2

R.T.: 5.626 min
Delta R.T.: 0.002 min
Response: 33429481
Conc: 536.89 ng/ml



#4 AR-1016-2


R.T.: 4.781 min
Delta R.T.: 0.000 min
Response: 47178505
Conc: 573.62 ng/ml



#5 AR-1016-3


R.T.: 5.688 min
Delta R.T.: 0.001 min
Response: 21550668
Conc: 538.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.958 min
Delta R.T.: 0.000 min
Response: 24810660
Conc: 564.20 ng/ml




#6 AR-1016-4

R.T.: 5.787 min
Delta R.T.: 0.000 min
Response: 17462766
Conc: 541.07 ng/ml



#6 AR-1016-4


R.T.: 5.001 min
Delta R.T.: 0.000 min
Response: 20403149
Conc: 524.70 ng/ml



#7 AR-1016-5


R.T.: 6.083 min
Delta R.T.: 0.000 min
Response: 18148958
Conc: 532.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.214 min
Delta R.T.: 0.000 min
Response: 27282102
Conc: 551.53 ng/ml




#31 AR-1260-1

R.T.: 7.216 min
Delta R.T.: 0.000 min
Response: 27845799
Conc: 480.14 ng/ml



#31 AR-1260-1


R.T.: 6.254 min
Delta R.T.: -0.001 min
Response: 49735884
Conc: 502.52 ng/ml



#32 AR-1260-2


R.T.: 7.473 min
Delta R.T.: 0.000 min
Response: 30652889
Conc: 505.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.444 min
Delta R.T.: 0.000 min
Response: 58678978
Conc: 516.12 ng/ml




#33 AR-1260-3

R.T.: 7.834 min
Delta R.T.: 0.000 min
Response: 22638557
Conc: 511.33 ng/ml



#33 AR-1260-3


R.T.: 6.598 min
Delta R.T.: -0.001 min
Response: 52659086
Conc: 481.13 ng/ml



#34 AR-1260-4


R.T.: 8.059 min
Delta R.T.: 0.000 min
Response: 25280173
Conc: 512.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.072 min
Delta R.T.: -0.001 min
Response: 35996931
Conc: 423.33 ng/ml



#35 AR-1260-5

R.T.: 8.382 min
Delta R.T.: 0.000 min
Response: 47753221
Conc: 535.59 ng/ml



#35 AR-1260-5

R.T.: 7.314 min
Delta R.T.: -0.001 min
Response: 76192407
Conc: 449.30 ng/ml