

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072423\
 Data File : PP058927.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2023 23:17
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 25 05:07:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 22 02:26:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.425	3.665	70569990	107.4E6	51.261	55.822
2) SA Decachlor...	10.223	8.708	50535662	74157166	58.661	52.410
Target Compounds						
26) L6 AR-1254-1	6.461	5.565	26920007	53556879	507.518	529.593
27) L6 AR-1254-2	6.681	5.714	39931325	46611878	510.019	539.137
28) L6 AR-1254-3	7.048	6.119	40536975	73513656	517.232	535.094
29) L6 AR-1254-4	7.334	6.349	26782668	41890021	512.946	524.277
30) L6 AR-1254-5	7.755	6.768	28710474	62735086	533.942	521.115



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072423\
Data File : PP058927.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2023 23:17
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 25 05:07:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 02:26:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.425 min
Delta R.T.: 0.000 min
Response: 70569990
Conc: 51.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.665 min
Delta R.T.: 0.000 min
Response: 107351754
Conc: 55.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.223 min
Delta R.T.: -0.004 min
Response: 50535662
Conc: 58.66 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.708 min
Delta R.T.: -0.006 min
Response: 74157166
Conc: 52.41 ng/ml



#26 AR-1254-1


R.T.: 6.461 min
Delta R.T.: -0.002 min
Response: 26920007
Conc: 507.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.565 min
Delta R.T.: -0.005 min
Response: 53556879
Conc: 529.59 ng/ml




#27 AR-1254-2

R.T.: 6.681 min
Delta R.T.: -0.002 min
Response: 39931325
Conc: 510.02 ng/ml



#27 AR-1254-2


R.T.: 5.714 min
Delta R.T.: -0.005 min
Response: 46611878
Conc: 539.14 ng/ml



#28 AR-1254-3


R.T.: 7.048 min
Delta R.T.: -0.002 min
Response: 40536975
Conc: 517.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.119 min
Delta R.T.: -0.005 min
Response: 73513656
Conc: 535.09 ng/ml




#29 AR-1254-4

R.T.: 7.334 min
Delta R.T.: -0.002 min
Response: 26782668
Conc: 512.95 ng/ml



#29 AR-1254-4


R.T.: 6.349 min
Delta R.T.: -0.005 min
Response: 41890021
Conc: 524.28 ng/ml



#30 AR-1254-5

R.T.: 7.755 min
Delta R.T.: -0.001 min
Response: 28710474
Conc: 533.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.768 min
Delta R.T.: -0.006 min
Response: 62735086
Conc: 521.12 ng/ml