

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072522\
 Data File : PP049806.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2022 00:15
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 03:56:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.351	3.600	78645334	141.8E6	54.196	58.004
2) SA Decachlor...	10.099	8.564	73554660	58552380	52.004	50.277
Target Compounds						
26) L6 AR-1254-1	6.384	5.467	31652978	59007146	502.578	465.243
27) L6 AR-1254-2	6.602	5.613	48093923	50733317	493.664	458.996
28) L6 AR-1254-3	6.972	6.016	51121342	73877386	494.771	453.864
29) L6 AR-1254-4	7.260	6.243	39416991	43494632	492.380	449.524
30) L6 AR-1254-5	7.680	6.659	42961850	60217679	483.358	454.420



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072522\
Data File : PP049806.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2022 00:15
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 26 03:56:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.351 min
Delta R.T.: -0.002 min
Response: 78645334
Conc: 54.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.600 min
Delta R.T.: -0.002 min
Response: 141781958
Conc: 58.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.099 min
Delta R.T.: -0.008 min
Response: 73554660
Conc: 52.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.564 min
Delta R.T.: -0.009 min
Response: 58552380
Conc: 50.28 ng/ml



#26 AR-1254-1


R.T.: 6.384 min
Delta R.T.: -0.002 min
Response: 31652978
Conc: 502.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.467 min
Delta R.T.: -0.004 min
Response: 59007146
Conc: 465.24 ng/ml




#27 AR-1254-2

R.T.: 6.602 min
Delta R.T.: -0.002 min
Response: 48093923
Conc: 493.66 ng/ml



#27 AR-1254-2


R.T.: 5.613 min
Delta R.T.: -0.003 min
Response: 50733317
Conc: 459.00 ng/ml



#28 AR-1254-3


R.T.: 6.972 min
Delta R.T.: -0.002 min
Response: 51121342
Conc: 494.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.016 min
Delta R.T.: -0.004 min
Response: 73877386
Conc: 453.86 ng/ml




#29 AR-1254-4

R.T.: 7.260 min
Delta R.T.: -0.001 min
Response: 39416991
Conc: 492.38 ng/ml



#29 AR-1254-4


R.T.: 6.243 min
Delta R.T.: -0.004 min
Response: 43494632
Conc: 449.52 ng/ml



#30 AR-1254-5

R.T.: 7.680 min
Delta R.T.: 0.000 min
Response: 42961850
Conc: 483.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.659 min
Delta R.T.: -0.004 min
Response: 60217679
Conc: 454.42 ng/ml