

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072522\
 Data File : PP049817.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2022 04:01
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 07:25:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.599	78841740	139.5E6	54.331	57.091
2) SA Decachlor...	10.097	8.562	74763012	60210728	52.858	51.701
Target Compounds						
16) L4 AR-1242-1	5.525	4.670	23063797	35635019	515.864	508.066
17) L4 AR-1242-2	5.548	4.688	33191243	50023828	526.426	506.951
18) L4 AR-1242-3	5.609	4.863	20383580	26646613	524.262	510.605
19) L4 AR-1242-4	5.710	4.945	17112582	25919441	530.452	494.534
20) L4 AR-1242-5	6.449	5.465	18196232	31380128	529.093	493.670



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072522\
Data File : PP049817.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2022 04:01
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 26 07:25:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.349 min
Delta R.T.: -0.004 min
Response: 78841740
Conc: 54.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.599 min
Delta R.T.: -0.003 min
Response: 139549507
Conc: 57.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.097 min
Delta R.T.: -0.010 min
Response: 74763012
Conc: 52.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.562 min
Delta R.T.: -0.011 min
Response: 60210728
Conc: 51.70 ng/ml



#16 AR-1242-1


R.T.: 5.525 min
Delta R.T.: -0.004 min
Response: 23063797
Conc: 515.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.670 min
Delta R.T.: -0.005 min
Response: 35635019
Conc: 508.07 ng/ml




#17 AR-1242-2

R.T.: 5.548 min
Delta R.T.: -0.004 min
Response: 33191243
Conc: 526.43 ng/ml



#17 AR-1242-2


R.T.: 4.688 min
Delta R.T.: -0.006 min
Response: 50023828
Conc: 506.95 ng/ml



#18 AR-1242-3


R.T.: 5.609 min
Delta R.T.: -0.004 min
Response: 20383580
Conc: 524.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.863 min
Delta R.T.: -0.005 min
Response: 26646613
Conc: 510.61 ng/ml




#19 AR-1242-4

R.T.: 5.710 min
Delta R.T.: -0.004 min
Response: 17112582
Conc: 530.45 ng/ml



#19 AR-1242-4


R.T.: 4.945 min
Delta R.T.: -0.006 min
Response: 25919441
Conc: 494.53 ng/ml



#20 AR-1242-5

R.T.: 6.449 min
Delta R.T.: -0.004 min
Response: 18196232
Conc: 529.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.465 min
Delta R.T.: -0.006 min
Response: 31380128
Conc: 493.67 ng/ml