

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072523\
 Data File : PP058947.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jul 2023 15:54
 Operator : YP\AJ
 Sample : 03741-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 25 17:16:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 22 02:26:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.425	3.664	27026519	38333096	19.631	19.933
2)	SA Decachlor...	10.219	8.707	17264786	26163402	20.041	18.491
Target Compounds							
3)	L1 AR-1016-1	5.593	4.760	264654	116203	6.061	2.022 #
4)	L1 AR-1016-2	5.625	4.787	460241	119583	7.392	1.454 #
5)	L1 AR-1016-3	5.683	4.959	214026	147712	5.346	3.359 #
6)	L1 AR-1016-4	5.802f	5.003	82934	52186	2.570	1.342 #
7)	L1 AR-1016-5	6.079	5.202	36800	263923	1.079	5.335 #
8)	L2 AR-1221-1	4.631	0.000	418135	0	23.512	N.D. #
9)	L2 AR-1221-2	4.733	3.966	440571	759952	33.609	42.611 #
10)	L2 AR-1221-3	4.810	4.036	41156	115743	1.061	2.098 #
11)	L3 AR-1232-1	4.810	4.036	41156	115743	1.280	2.550 #
12)	L3 AR-1232-2	5.323	4.787	40066	119583	2.212	3.097 #
13)	L3 AR-1232-3	5.625	4.959	460241	147712	15.534	7.110 #
14)	L3 AR-1232-4	5.802f	5.042	82934	32399	5.525	1.601 #
15)	L3 AR-1232-5	5.878	5.202	113762	263923	8.782	11.805 #
16)	L4 AR-1242-1	5.593	4.760	264654	116203	7.001	2.338 #
17)	L4 AR-1242-2	5.625	4.787	460241	119583	8.547	1.678 #
18)	L4 AR-1242-3	5.683	4.959	214026	147712	6.180	3.896 #
19)	L4 AR-1242-4	5.802f	5.042	82934	32399	2.946	0.805 #
20)	L4 AR-1242-5	6.533	5.565	97287	716593	3.428	14.886 #
21)	L5 AR-1248-1	5.593	4.760	264654	116203	9.564	3.226 #
22)	L5 AR-1248-2	5.878	5.003	113762	52186	2.646	0.921 #
23)	L5 AR-1248-3	6.079	5.042	36800	32399	0.785	0.566 #
24)	L5 AR-1248-4	6.498	5.202	103287	263923	2.184	3.883 #
25)	L5 AR-1248-5	6.533	5.610	97287	243946	2.123	4.003 #
26)	L6 AR-1254-1	6.461	5.565	160551	716593	3.027	7.086 #
27)	L6 AR-1254-2	6.679	5.713	579494	713576	7.402	8.254
28)	L6 AR-1254-3	7.047	6.119	803329	1834849	10.250	13.356 #
29)	L6 AR-1254-4	7.333	6.348	319469	693114	6.119	8.675 #
30)	L6 AR-1254-5	7.772f	6.768	2404572	4175766	44.719	34.686
31)	L7 AR-1260-1	7.211	6.249	599958	1190532	10.345	12.029
32)	L7 AR-1260-2	7.468	6.438	796334	1116206	13.142	9.818 #
33)	L7 AR-1260-3	7.833	6.591	166385	687168	3.758	6.279 #
34)	L7 AR-1260-4	8.077	7.063	90570	443170	1.836	5.212 #
35)	L7 AR-1260-5	8.381	7.308	272076	261544	3.052	1.542 #
36)	L8 AR-1262-1	7.833	6.858	166385	151192	2.438	2.767
37)	L8 AR-1262-2	8.381	7.063	272076	443170	2.482	3.843 #
38)	L8 AR-1262-3	8.693	7.594	564757	766260	7.206	8.706
39)	L8 AR-1262-4	8.783	7.657	522416	1178671	8.641	7.487
40)	L8 AR-1262-5	9.454	8.157	109758	193190	2.826	2.835
41)	L9 AR-1268-1	8.693	7.594	564757	766260	4.216	3.160 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072523\
 Data File : PP058947.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jul 2023 15:54
 Operator : YP\AJ
 Sample : 03741-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 25 17:16:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 22 02:26:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.783	7.657	522416	1178671	4.388	5.314
43) L9	AR-1268-3	9.018	7.867	546677	1082976	5.109	5.314
44) L9	AR-1268-4	9.454	8.157	109758	193190	2.640	2.613
45) L9	AR-1268-5	9.873	8.449	1680726	2827451	5.249	5.134



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072523\
Data File : PP058947.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jul 2023 15:54
Operator : YP\AJ
Sample : 03741-06
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 25 17:16:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 02:26:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.425 min
Delta R.T.: 0.000 min
Response: 27026519
Conc: 19.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.664 min
Delta R.T.: 0.000 min
Response: 38333096
Conc: 19.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.219 min
Delta R.T.: -0.007 min
Response: 17264786
Conc: 20.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.707 min
Delta R.T.: -0.008 min
Response: 26163402
Conc: 18.49 ng/ml



#3 AR-1016-1


R.T.: 5.593 min
Delta R.T.: -0.009 min
Response: 264654
Conc: 6.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.760 min
Delta R.T.: -0.002 min
Response: 116203
Conc: 2.02 ng/ml




#4 AR-1016-2

R.T.: 5.625 min
Delta R.T.: 0.001 min
Response: 460241
Conc: 7.39 ng/ml



#4 AR-1016-2


R.T.: 4.787 min
Delta R.T.: 0.006 min
Response: 119583
Conc: 1.45 ng/ml



#5 AR-1016-3


R.T.: 5.683 min
Delta R.T.: -0.004 min
Response: 214026
Conc: 5.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.959 min
Delta R.T.: 0.001 min
Response: 147712
Conc: 3.36 ng/ml




#6 AR-1016-4

R.T.: 5.802 min
Delta R.T.: 0.016 min
Response: 82934
Conc: 2.57 ng/ml



#6 AR-1016-4


R.T.: 5.003 min
Delta R.T.: 0.003 min
Response: 52186
Conc: 1.34 ng/ml



#7 AR-1016-5


R.T.: 6.079 min
Delta R.T.: -0.004 min
Response: 36800
Conc: 1.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.202 min
Delta R.T.: -0.013 min
Response: 263923
Conc: 5.34 ng/ml




#8 AR-1221-1

R.T.: 4.631 min
Delta R.T.: -0.003 min
Response: 418135
Conc: 23.51 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.881 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.733 min
Delta R.T.: 0.012 min
Response: 440571
Conc: 33.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.966 min
Delta R.T.: -0.002 min
Response: 759952
Conc: 42.61 ng/ml




#10 AR-1221-3

R.T.: 4.810 min
Delta R.T.: 0.012 min
Response: 41156
Conc: 1.06 ng/ml



#10 AR-1221-3


R.T.: 4.036 min
Delta R.T.: -0.009 min
Response: 115743
Conc: 2.10 ng/ml



#11 AR-1232-1


R.T.: 4.810 min
Delta R.T.: 0.011 min
Response: 41156
Conc: 1.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.036 min
Delta R.T.: -0.009 min
Response: 115743
Conc: 2.55 ng/ml




#12 AR-1232-2

R.T.: 5.323 min
Delta R.T.: -0.009 min
Response: 40066
Conc: 2.21 ng/ml



#12 AR-1232-2


R.T.: 4.787 min
Delta R.T.: 0.006 min
Response: 119583
Conc: 3.10 ng/ml



#13 AR-1232-3


R.T.: 5.625 min
Delta R.T.: 0.001 min
Response: 460241
Conc: 15.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.959 min
Delta R.T.: 0.001 min
Response: 147712
Conc: 7.11 ng/ml




#14 AR-1232-4

R.T.: 5.802 min
Delta R.T.: 0.016 min
Response: 82934
Conc: 5.53 ng/ml



#14 AR-1232-4


R.T.: 5.042 min
Delta R.T.: 0.000 min
Response: 32399
Conc: 1.60 ng/ml



#15 AR-1232-5


R.T.: 5.878 min
Delta R.T.: 0.000 min
Response: 113762
Conc: 8.78 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.202 min
Delta R.T.: -0.014 min
Response: 263923
Conc: 11.80 ng/ml




#16 AR-1242-1

R.T.: 5.593 min
Delta R.T.: -0.009 min
Response: 264654
Conc: 7.00 ng/ml



#16 AR-1242-1


R.T.: 4.760 min
Delta R.T.: -0.002 min
Response: 116203
Conc: 2.34 ng/ml



#17 AR-1242-2


R.T.: 5.625 min
Delta R.T.: 0.000 min
Response: 460241
Conc: 8.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.787 min
Delta R.T.: 0.006 min
Response: 119583
Conc: 1.68 ng/ml




#18 AR-1242-3

R.T.: 5.683 min
Delta R.T.: -0.005 min
Response: 214026
Conc: 6.18 ng/ml



#18 AR-1242-3


R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 147712
Conc: 3.90 ng/ml



#19 AR-1242-4


R.T.: 5.802 min
Delta R.T.: 0.016 min
Response: 82934
Conc: 2.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.042 min
Delta R.T.: 0.000 min
Response: 32399
Conc: 0.80 ng/ml




#20 AR-1242-5

R.T.: 6.533 min
Delta R.T.: 0.006 min
Response: 97287
Conc: 3.43 ng/ml



#20 AR-1242-5


R.T.: 5.565 min
Delta R.T.: -0.005 min
Response: 716593
Conc: 14.89 ng/ml



#21 AR-1248-1


R.T.: 5.593 min
Delta R.T.: -0.010 min
Response: 264654
Conc: 9.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.760 min
Delta R.T.: -0.003 min
Response: 116203
Conc: 3.23 ng/ml




#22 AR-1248-2

R.T.: 5.878 min
Delta R.T.: 0.000 min
Response: 113762
Conc: 2.65 ng/ml



#22 AR-1248-2


R.T.: 5.003 min
Delta R.T.: 0.003 min
Response: 52186
Conc: 0.92 ng/ml



#23 AR-1248-3


R.T.: 6.079 min
Delta R.T.: -0.005 min
Response: 36800
Conc: 0.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.042 min
Delta R.T.: 0.000 min
Response: 32399
Conc: 0.57 ng/ml




#24 AR-1248-4

R.T.: 6.498 min
Delta R.T.: 0.009 min
Response: 103287
Conc: 2.18 ng/ml



#24 AR-1248-4


R.T.: 5.202 min
Delta R.T.: -0.014 min
Response: 263923
Conc: 3.88 ng/ml



#25 AR-1248-5


R.T.: 6.533 min
Delta R.T.: 0.005 min
Response: 97287
Conc: 2.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.610 min
Delta R.T.: -0.001 min
Response: 243946
Conc: 4.00 ng/ml




#26 AR-1254-1

R.T.: 6.461 min
Delta R.T.: -0.002 min
Response: 160551
Conc: 3.03 ng/ml



#26 AR-1254-1


R.T.: 5.565 min
Delta R.T.: -0.005 min
Response: 716593
Conc: 7.09 ng/ml



#27 AR-1254-2


R.T.: 6.679 min
Delta R.T.: -0.004 min
Response: 579494
Conc: 7.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.713 min
Delta R.T.: -0.005 min
Response: 713576
Conc: 8.25 ng/ml




#28 AR-1254-3

R.T.: 7.047 min
Delta R.T.: -0.003 min
Response: 803329
Conc: 10.25 ng/ml



#28 AR-1254-3


R.T.: 6.119 min
Delta R.T.: -0.005 min
Response: 1834849
Conc: 13.36 ng/ml



#29 AR-1254-4


R.T.: 7.333 min
Delta R.T.: -0.003 min
Response: 319469
Conc: 6.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.348 min
Delta R.T.: -0.006 min
Response: 693114
Conc: 8.67 ng/ml




#30 AR-1254-5

R.T.: 7.772 min
Delta R.T.: 0.015 min
Response: 2404572
Conc: 44.72 ng/ml



#30 AR-1254-5


R.T.: 6.768 min
Delta R.T.: -0.007 min
Response: 4175766
Conc: 34.69 ng/ml



#31 AR-1260-1


R.T.: 7.211 min
Delta R.T.: -0.004 min
Response: 599958
Conc: 10.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.249 min
Delta R.T.: -0.006 min
Response: 1190532
Conc: 12.03 ng/ml




#32 AR-1260-2

R.T.: 7.468 min
Delta R.T.: -0.004 min
Response: 796334
Conc: 13.14 ng/ml



#32 AR-1260-2


R.T.: 6.438 min
Delta R.T.: -0.006 min
Response: 1116206
Conc: 9.82 ng/ml



#33 AR-1260-3


R.T.: 7.833 min
Delta R.T.: 0.000 min
Response: 166385
Conc: 3.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.591 min
Delta R.T.: -0.008 min
Response: 687168
Conc: 6.28 ng/ml




#34 AR-1260-4

R.T.: 8.077 min
Delta R.T.: 0.018 min
Response: 90570
Conc: 1.84 ng/ml



#34 AR-1260-4


R.T.: 7.063 min
Delta R.T.: -0.010 min
Response: 443170
Conc: 5.21 ng/ml



#35 AR-1260-5


R.T.: 8.381 min
Delta R.T.: 0.000 min
Response: 272076
Conc: 3.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.308 min
Delta R.T.: -0.008 min
Response: 261544
Conc: 1.54 ng/ml




#36 AR-1262-1

R.T.: 7.833 min
Delta R.T.: 0.000 min
Response: 166385
Conc: 2.44 ng/ml



#36 AR-1262-1


R.T.: 6.858 min
Delta R.T.: -0.007 min
Response: 151192
Conc: 2.77 ng/ml



#37 AR-1262-2


R.T.: 8.381 min
Delta R.T.: 0.000 min
Response: 272076
Conc: 2.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.063 min
Delta R.T.: -0.009 min
Response: 443170
Conc: 3.84 ng/ml




#38 AR-1262-3

R.T.: 8.693 min
Delta R.T.: -0.005 min
Response: 564757
Conc: 7.21 ng/ml



#38 AR-1262-3


R.T.: 7.594 min
Delta R.T.: -0.007 min
Response: 766260
Conc: 8.71 ng/ml



#39 AR-1262-4


R.T.: 8.783 min
Delta R.T.: -0.006 min
Response: 522416
Conc: 8.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.657 min
Delta R.T.: -0.007 min
Response: 1178671
Conc: 7.49 ng/ml




#40 AR-1262-5

R.T.: 9.454 min
Delta R.T.: 0.000 min
Response: 109758
Conc: 2.83 ng/ml



#40 AR-1262-5


R.T.: 8.157 min
Delta R.T.: -0.004 min
Response: 193190
Conc: 2.83 ng/ml



#41 AR-1268-1


R.T.: 8.693 min
Delta R.T.: -0.003 min
Response: 564757
Conc: 4.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.594 min
Delta R.T.: -0.007 min
Response: 766260
Conc: 3.16 ng/ml




#42 AR-1268-2

R.T.: 8.783 min
Delta R.T.: -0.008 min
Response: 522416
Conc: 4.39 ng/ml



#42 AR-1268-2


R.T.: 7.657 min
Delta R.T.: -0.009 min
Response: 1178671
Conc: 5.31 ng/ml



#43 AR-1268-3


R.T.: 9.018 min
Delta R.T.: -0.005 min
Response: 546677
Conc: 5.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.867 min
Delta R.T.: -0.008 min
Response: 1082976
Conc: 5.31 ng/ml




#44 AR-1268-4

R.T.: 9.454 min
Delta R.T.: 0.000 min
Response: 109758
Conc: 2.64 ng/ml



#44 AR-1268-4


R.T.: 8.157 min
Delta R.T.: -0.005 min
Response: 193190
Conc: 2.61 ng/ml



#45 AR-1268-5

R.T.: 9.873 min
Delta R.T.: -0.005 min
Response: 1680726
Conc: 5.25 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.449 min
Delta R.T.: -0.008 min
Response: 2827451
Conc: 5.13 ng/ml