

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072622\
 Data File : PP049850.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2022 17:53
 Operator : YP\AJ
 Sample : N3839-08
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 FIELD-BLANK

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/27/2022
 Supervised By :Ankita Jodhani 07/27/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 00:35:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.350	3.601	27087624	48350504	18.666	19.781
2) SA Decachlor...	10.097	8.563	22827851	19072912	16.139m	16.377

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072622\
Data File : PP049850.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2022 17:53
Operator : YP\AJ
Sample : N3839-08
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
FIELD-BLANK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/27/2022
Supervised By :Ankita Jodhani 07/27/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 27 00:35:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.350 min
Delta R.T.: -0.003 min
Response: 27087624
Conc: 18.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
FIELD-BLANK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/27/2022
Supervised By :Ankita Jodhani 07/27/2022




#1 Tetrachloro-m-xylene

R.T.: 3.601 min
Delta R.T.: -0.001 min
Response: 48350504
Conc: 19.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.097 min
Delta R.T.: -0.010 min
Response: 22827851
Conc: 16.14 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.563 min
Delta R.T.: -0.009 min
Response: 19072912
Conc: 16.38 ng/ml