

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072722\
 Data File : PP049907.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2022 00:45
 Operator : YP\AJ
 Sample : N3905-08 10X
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BCZ74271-(1-2)

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/28/2022
 Supervised By :Ankita Jodhani 07/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 28 05:48:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.350	3.599	2255016	3943301	1.554	1.613
2) SA Decachlor...	10.104	8.567	2138832	1759421	1.512	1.511
Target Compounds						
26) L6 AR-1254-1	6.383	5.467	985608	2482958	15.649m	19.577 #
27) L6 AR-1254-2	6.603	5.612	1401607	1613269	14.387	14.596
28) L6 AR-1254-3	6.972	6.015	1274643	2344507	12.336m	14.403
29) L6 AR-1254-4	7.259	6.242	1407997	1542002	17.588m	15.937
30) L6 AR-1254-5	7.679	6.658	917043	1732864	10.318m	13.077 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072722\
Data File : PP049907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2022 00:45
Operator : YP\AJ
Sample : N3905-08 10X
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
BCZ74271-(1-2)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/28/2022
Supervised By :Ankita Jodhani 07/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 05:48:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.350 min
Delta R.T.: -0.003 min
Response: 2255016
Conc: 1.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
BCZ74271-(1-2)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/28/2022
Supervised By :Ankita Jodhani 07/28/2022




#1 Tetrachloro-m-xylene

R.T.: 3.599 min
Delta R.T.: -0.003 min
Response: 3943301
Conc: 1.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.104 min
Delta R.T.: -0.003 min
Response: 2138832
Conc: 1.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.567 min
Delta R.T.: -0.006 min
Response: 1759421
Conc: 1.51 ng/ml




#26 AR-1254-1

R.T.: 6.383 min
Delta R.T.: -0.003 min
Response: 985608
Conc: 15.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
BCZ74271-(1-2)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/28/2022
Supervised By :Ankita Jodhani 07/28/2022




#26 AR-1254-1

R.T.: 5.467 min
Delta R.T.: -0.004 min
Response: 2482958
Conc: 19.58 ng/ml




#27 AR-1254-2

R.T.: 6.603 min
Delta R.T.: -0.002 min
Response: 1401607
Conc: 14.39 ng/ml



#27 AR-1254-2

R.T.: 5.612 min
Delta R.T.: -0.005 min
Response: 1613269
Conc: 14.60 ng/ml




#28 AR-1254-3

R.T.: 6.972 min
Delta R.T.: -0.002 min
Response: 1274643
Conc: 12.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
BCZ74271-(1-2)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/28/2022
Supervised By :Ankita Jodhani 07/28/2022




#28 AR-1254-3

R.T.: 6.015 min
Delta R.T.: -0.005 min
Response: 2344507
Conc: 14.40 ng/ml




#29 AR-1254-4

R.T.: 7.259 min
Delta R.T.: -0.002 min
Response: 1407997
Conc: 17.59 ng/ml m



#29 AR-1254-4

R.T.: 6.242 min
Delta R.T.: -0.005 min
Response: 1542002
Conc: 15.94 ng/ml




#30 AR-1254-5

R.T.: 7.679 min
Delta R.T.: -0.001 min
Response: 917043
Conc: 10.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
BCZ74271-(1-2)

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/28/2022
Supervised By :Ankita Jodhani 07/28/2022



#30 AR-1254-5

R.T.: 6.658 min
Delta R.T.: -0.005 min
Response: 1732864
Conc: 13.08 ng/ml