

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072723\
 Data File : PP059024.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jul 2023 16:36
 Operator : YP\AJ
 Sample : 03787-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 21:41:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 22 02:26:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.425	3.663	25599407	35757086	18.595	18.594
2)	SA Decachlor...	10.223	8.705	23673868	37348620	27.480	26.396
Target Compounds							
3)	L1 AR-1016-1	5.582f	4.769	79373	73466	1.818	1.278 #
4)	L1 AR-1016-2	5.648f	4.777	82856	53818	1.331	0.654 #
5)	L1 AR-1016-3	5.683	4.959	85627	257989	2.139	5.867 #
6)	L1 AR-1016-4	5.797	4.998	8688	26631	0.269	0.685 #
7)	L1 AR-1016-5	6.075	5.200	126135	279192	3.698	5.644 #
8)	L2 AR-1221-1	4.634	3.901f	269423	14785	15.150	0.626 #
9)	L2 AR-1221-2	4.732	3.965	463353	676771	35.347	37.947
10)	L2 AR-1221-3	4.793	4.043	27505	235080	0.709	4.261 #
11)	L3 AR-1232-1	4.793	4.043	27505	235080	0.855	5.179 #
12)	L3 AR-1232-2	5.328	4.777	5486	53818	0.303	1.394 #
13)	L3 AR-1232-3	5.648f	4.959	82856	257989	2.796	12.418 #
14)	L3 AR-1232-4	5.797	5.048	8688	75735	0.579	3.742 #
15)	L3 AR-1232-5	5.881	5.200f	52895	279192	4.083	12.488 #
16)	L4 AR-1242-1	5.582f	4.769	79373	73466	2.100	1.478 #
17)	L4 AR-1242-2	5.648f	4.777	82856	53818	1.539	0.755 #
18)	L4 AR-1242-3	5.683	4.959	85627	257989	2.472	6.804 #
19)	L4 AR-1242-4	5.797	5.048	8688	75735	0.309	1.881 #
20)	L4 AR-1242-5	6.517	5.563	25205	302197	0.888	6.277 #
21)	L5 AR-1248-1	5.582f	4.769	79373	73466	2.868	2.039 #
22)	L5 AR-1248-2	5.881	4.998	52895	26631	1.230	0.470 #
23)	L5 AR-1248-3	6.075	5.048	126135	75735	2.692	1.322 #
24)	L5 AR-1248-4	6.486	5.200f	33022	279192	0.698	4.108 #
25)	L5 AR-1248-5	6.539	5.608	36625	114847	0.799	1.885 #
26)	L6 AR-1254-1	6.459	5.563	-2282	302197	N.D.	2.988 #
27)	L6 AR-1254-2	6.710f	5.718	276496	85125	3.532	0.985 #
28)	L6 AR-1254-3	7.051	6.143f	270821	463486	3.456	3.374
29)	L6 AR-1254-4	7.334	6.339	62827	60783	1.203	0.761 #
30)	L6 AR-1254-5	0.000	6.767	0	180082	N.D.	1.496 #
31)	L7 AR-1260-1	7.206	6.263	78743	38966	1.358	0.394 #
32)	L7 AR-1260-2	7.475	6.441	141115	17713	2.329	0.156 #
33)	L7 AR-1260-3	7.825	6.598	27084	175262	0.612	1.601 #
34)	L7 AR-1260-4	8.067	7.060	41422	375051	0.840	4.411 #
35)	L7 AR-1260-5	8.372	7.314	27889	99762	0.313	0.588 #
36)	L8 AR-1262-1	7.825	6.855	27084	66097	0.397	1.209 #
37)	L8 AR-1262-2	8.372	7.060	27889	375051	0.254	3.252 #
39)	L8 AR-1262-4	8.788	7.652	502773	647826	8.316	4.115 #
40)	L8 AR-1262-5	9.428f	8.157	25891	34032	0.667	0.499 #
41)	L9 AR-1268-1	8.692	7.595	57551	-79622	0.430	N.D. #
42)	L9 AR-1268-2	8.788	7.652	502773	647826	4.223	2.921 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072723\
 Data File : PP059024.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jul 2023 16:36
 Operator : YP\AJ
 Sample : 03787-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 21:41:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 22 02:26:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43)	L9 AR-1268-3	9.022	7.869	25930	209216	0.242	1.027 #
44)	L9 AR-1268-4	9.428f	8.157	25891	34032	0.623	0.460 #
45)	L9 AR-1268-5	9.885	8.450	50178	469007	0.157	0.852 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072723\
Data File : PP059024.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2023 16:36
Operator : YP\AJ
Sample : 03787-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 27 21:41:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 02:26:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.425 min
Delta R.T.: 0.000 min
Response: 25599407
Conc: 18.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.663 min
Delta R.T.: -0.001 min
Response: 35757086
Conc: 18.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.223 min
Delta R.T.: -0.003 min
Response: 23673868
Conc: 27.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.705 min
Delta R.T.: -0.009 min
Response: 37348620
Conc: 26.40 ng/ml



#3 AR-1016-1


R.T.: 5.582 min
Delta R.T.: -0.020 min
Response: 79373
Conc: 1.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.769 min
Delta R.T.: 0.007 min
Response: 73466
Conc: 1.28 ng/ml




#4 AR-1016-2

R.T.: 5.648 min
Delta R.T.: 0.024 min
Response: 82856
Conc: 1.33 ng/ml



#4 AR-1016-2


R.T.: 4.777 min
Delta R.T.: -0.003 min
Response: 53818
Conc: 0.65 ng/ml



#5 AR-1016-3


R.T.: 5.683 min
Delta R.T.: -0.004 min
Response: 85627
Conc: 2.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 257989
Conc: 5.87 ng/ml




#6 AR-1016-4

R.T.: 5.797 min
Delta R.T.: 0.011 min
Response: 8688
Conc: 0.27 ng/ml



#6 AR-1016-4


R.T.: 4.998 min
Delta R.T.: -0.003 min
Response: 26631
Conc: 0.68 ng/ml



#7 AR-1016-5


R.T.: 6.075 min
Delta R.T.: -0.007 min
Response: 126135
Conc: 3.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.200 min
Delta R.T.: -0.015 min
Response: 279192
Conc: 5.64 ng/ml




#8 AR-1221-1

R.T.: 4.634 min
Delta R.T.: 0.000 min
Response: 269423
Conc: 15.15 ng/ml



#8 AR-1221-1


R.T.: 3.901 min
Delta R.T.: 0.019 min
Response: 14785
Conc: 0.63 ng/ml



#9 AR-1221-2


R.T.: 4.732 min
Delta R.T.: 0.012 min
Response: 463353
Conc: 35.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.965 min
Delta R.T.: -0.003 min
Response: 676771
Conc: 37.95 ng/ml




#10 AR-1221-3

R.T.: 4.793 min
Delta R.T.: -0.005 min
Response: 27505
Conc: 0.71 ng/ml



#10 AR-1221-3


R.T.: 4.043 min
Delta R.T.: -0.001 min
Response: 235080
Conc: 4.26 ng/ml



#11 AR-1232-1


R.T.: 4.793 min
Delta R.T.: -0.005 min
Response: 27505
Conc: 0.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.043 min
Delta R.T.: -0.002 min
Response: 235080
Conc: 5.18 ng/ml




#12 AR-1232-2

R.T.: 5.328 min
Delta R.T.: -0.004 min
Response: 5486
Conc: 0.30 ng/ml



#12 AR-1232-2


R.T.: 4.777 min
Delta R.T.: -0.004 min
Response: 53818
Conc: 1.39 ng/ml



#13 AR-1232-3


R.T.: 5.648 min
Delta R.T.: 0.024 min
Response: 82856
Conc: 2.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 257989
Conc: 12.42 ng/ml




#14 AR-1232-4

R.T.: 5.797 min
Delta R.T.: 0.011 min
Response: 8688
Conc: 0.58 ng/ml



#14 AR-1232-4


R.T.: 5.048 min
Delta R.T.: 0.005 min
Response: 75735
Conc: 3.74 ng/ml



#15 AR-1232-5


R.T.: 5.881 min
Delta R.T.: 0.003 min
Response: 52895
Conc: 4.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.200 min
Delta R.T.: -0.016 min
Response: 279192
Conc: 12.49 ng/ml




#16 AR-1242-1

R.T.: 5.582 min
Delta R.T.: -0.020 min
Response: 79373
Conc: 2.10 ng/ml



#16 AR-1242-1


R.T.: 4.769 min
Delta R.T.: 0.007 min
Response: 73466
Conc: 1.48 ng/ml



#17 AR-1242-2


R.T.: 5.648 min
Delta R.T.: 0.023 min
Response: 82856
Conc: 1.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.777 min
Delta R.T.: -0.004 min
Response: 53818
Conc: 0.75 ng/ml




#18 AR-1242-3

R.T.: 5.683 min
Delta R.T.: -0.004 min
Response: 85627
Conc: 2.47 ng/ml



#18 AR-1242-3


R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 257989
Conc: 6.80 ng/ml



#19 AR-1242-4


R.T.: 5.797 min
Delta R.T.: 0.011 min
Response: 8688
Conc: 0.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.048 min
Delta R.T.: 0.005 min
Response: 75735
Conc: 1.88 ng/ml




#20 AR-1242-5

R.T.: 6.517 min
Delta R.T.: -0.010 min
Response: 25205
Conc: 0.89 ng/ml



#20 AR-1242-5


R.T.: 5.563 min
Delta R.T.: -0.007 min
Response: 302197
Conc: 6.28 ng/ml



#21 AR-1248-1


R.T.: 5.582 min
Delta R.T.: -0.021 min
Response: 79373
Conc: 2.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.769 min
Delta R.T.: 0.006 min
Response: 73466
Conc: 2.04 ng/ml




#22 AR-1248-2

R.T.: 5.881 min
Delta R.T.: 0.002 min
Response: 52895
Conc: 1.23 ng/ml



#22 AR-1248-2


R.T.: 4.998 min
Delta R.T.: -0.003 min
Response: 26631
Conc: 0.47 ng/ml



#23 AR-1248-3


R.T.: 6.075 min
Delta R.T.: -0.008 min
Response: 126135
Conc: 2.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.048 min
Delta R.T.: 0.005 min
Response: 75735
Conc: 1.32 ng/ml




#24 AR-1248-4

R.T.: 6.486 min
Delta R.T.: -0.003 min
Response: 33022
Conc: 0.70 ng/ml



#24 AR-1248-4


R.T.: 5.200 min
Delta R.T.: -0.016 min
Response: 279192
Conc: 4.11 ng/ml



#25 AR-1248-5


R.T.: 6.539 min
Delta R.T.: 0.011 min
Response: 36625
Conc: 0.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.608 min
Delta R.T.: -0.003 min
Response: 114847
Conc: 1.88 ng/ml




#26 AR-1254-1

R.T.: 6.459 min
Delta R.T.: -0.003 min
Response: -2282
Conc: N.D.



#26 AR-1254-1


R.T.: 5.563 min
Delta R.T.: -0.007 min
Response: 302197
Conc: 2.99 ng/ml



#27 AR-1254-2


R.T.: 6.710 min
Delta R.T.: 0.027 min
Response: 276496
Conc: 3.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.718 min
Delta R.T.: 0.000 min
Response: 85125
Conc: 0.98 ng/ml




#28 AR-1254-3

R.T.: 7.051 min
Delta R.T.: 0.000 min
Response: 270821
Conc: 3.46 ng/ml



#28 AR-1254-3


R.T.: 6.143 min
Delta R.T.: 0.018 min
Response: 463486
Conc: 3.37 ng/ml



#29 AR-1254-4


R.T.: 7.334 min
Delta R.T.: -0.002 min
Response: 62827
Conc: 1.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.339 min
Delta R.T.: -0.015 min
Response: 60783
Conc: 0.76 ng/ml




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.757 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 6.767 min
Delta R.T.: -0.007 min
Response: 180082
Conc: 1.50 ng/ml



#31 AR-1260-1


R.T.: 7.206 min
Delta R.T.: -0.009 min
Response: 78743
Conc: 1.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.263 min
Delta R.T.: 0.007 min
Response: 38966
Conc: 0.39 ng/ml




#32 AR-1260-2

R.T.: 7.475 min
Delta R.T.: 0.003 min
Response: 141115
Conc: 2.33 ng/ml



#32 AR-1260-2


R.T.: 6.441 min
Delta R.T.: -0.003 min
Response: 17713
Conc: 0.16 ng/ml



#33 AR-1260-3


R.T.: 7.825 min
Delta R.T.: -0.009 min
Response: 27084
Conc: 0.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.598 min
Delta R.T.: 0.000 min
Response: 175262
Conc: 1.60 ng/ml




#34 AR-1260-4

R.T.: 8.067 min
Delta R.T.: 0.008 min
Response: 41422
Conc: 0.84 ng/ml



#34 AR-1260-4


R.T.: 7.060 min
Delta R.T.: -0.012 min
Response: 375051
Conc: 4.41 ng/ml



#35 AR-1260-5


R.T.: 8.372 min
Delta R.T.: -0.009 min
Response: 27889
Conc: 0.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.314 min
Delta R.T.: -0.001 min
Response: 99762
Conc: 0.59 ng/ml




#36 AR-1262-1

R.T.: 7.825 min
Delta R.T.: -0.008 min
Response: 27084
Conc: 0.40 ng/ml



#36 AR-1262-1


R.T.: 6.855 min
Delta R.T.: -0.010 min
Response: 66097
Conc: 1.21 ng/ml



#37 AR-1262-2


R.T.: 8.372 min
Delta R.T.: -0.009 min
Response: 27889
Conc: 0.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.060 min
Delta R.T.: -0.011 min
Response: 375051
Conc: 3.25 ng/ml




#38 AR-1262-3

R.T.: 8.692 min
Delta R.T.: -0.005 min
Response: 57551
Conc: 0.73 ng/ml



#38 AR-1262-3


R.T.: 7.595 min
Delta R.T.: -0.006 min
Response: -79622
Conc: N.D.



#39 AR-1262-4


R.T.: 8.788 min
Delta R.T.: 0.000 min
Response: 502773
Conc: 8.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.652 min
Delta R.T.: -0.012 min
Response: 647826
Conc: 4.12 ng/ml




#40 AR-1262-5

R.T.: 9.428 min
Delta R.T.: -0.026 min
Response: 25891
Conc: 0.67 ng/ml



#40 AR-1262-5


R.T.: 8.157 min
Delta R.T.: -0.004 min
Response: 34032
Conc: 0.50 ng/ml



#41 AR-1268-1


R.T.: 8.692 min
Delta R.T.: -0.004 min
Response: 57551
Conc: 0.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.595 min
Delta R.T.: -0.007 min
Response: -79622
Conc: N.D.




#42 AR-1268-2

R.T.: 8.788 min
Delta R.T.: -0.002 min
Response: 502773
Conc: 4.22 ng/ml



#42 AR-1268-2


R.T.: 7.652 min
Delta R.T.: -0.014 min
Response: 647826
Conc: 2.92 ng/ml



#43 AR-1268-3


R.T.: 9.022 min
Delta R.T.: -0.001 min
Response: 25930
Conc: 0.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.869 min
Delta R.T.: -0.005 min
Response: 209216
Conc: 1.03 ng/ml




#44 AR-1268-4

R.T.: 9.428 min
Delta R.T.: -0.027 min
Response: 25891
Conc: 0.62 ng/ml



#44 AR-1268-4


R.T.: 8.157 min
Delta R.T.: -0.005 min
Response: 34032
Conc: 0.46 ng/ml



#45 AR-1268-5

R.T.: 9.885 min
Delta R.T.: 0.007 min
Response: 50178
Conc: 0.16 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.450 min
Delta R.T.: -0.007 min
Response: 469007
Conc: 0.85 ng/ml