

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072821\
 Data File : PP037758.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jul 2021 14:10
 Operator : DD\AJ
 Sample : M3203-03
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 NWB-1630

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 03:49:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.961	3.957	683509	404709	15.039	14.790
2)	SA Decachlor...	10.983	9.235	934473	441391	20.299	15.857
Target Compounds							
8)	L2 AR-1221-1	5.210	0.000	55069	0	103.831	N.D. #
9)	L2 AR-1221-2	5.313	0.000	28787	0	71.056	N.D. #
10)	L2 AR-1221-3	5.397	0.000	4892751	0	4015.588	N.D. #
11)	L3 AR-1232-1	5.397	0.000	4892751	0	4403.471	N.D. #
12)	L3 AR-1232-2	5.969f	0.000	40497	0	67.916	N.D. #
15)	L3 AR-1232-5	6.572	0.000	13184	0	34.280	N.D. #
22)	L5 AR-1248-2	6.572	0.000	13184	0	8.881	N.D. #
24)	L5 AR-1248-4	7.216	0.000	15745	0	8.330	N.D. #
26)	L6 AR-1254-1	7.200	0.000	24036	0	13.380	N.D. #
28)	L6 AR-1254-3	7.818f	0.000	75962	0	24.639	N.D. #
29)	L6 AR-1254-4	0.000	6.866f	0	10000	N.D.	7.371 #
31)	L7 AR-1260-1	7.974	6.775f	10854	18817	4.930	13.618 #
32)	L7 AR-1260-2	8.207f	0.000	12364	0	4.445	N.D. #
35)	L7 AR-1260-5	9.149	0.000	35869	0	7.773	N.D. #
37)	L8 AR-1262-2	9.149	0.000	35869	0	6.830	N.D. #
38)	L8 AR-1262-3	9.477	8.146f	228164	11993	59.548	8.518 #
41)	L9 AR-1268-1	9.477	8.146f	228164	11993	37.572	3.009 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072821\
Data File : PP037758.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jul 2021 14:10
Operator : DD\AJ
Sample : M3203-03
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
NWB-1630

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 03:49:28 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.961 min
Delta R.T.: 0.000 min
Response: 683509
Conc: 15.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
NWB-1630




#1 Tetrachloro-m-xylene

R.T.: 3.957 min
Delta R.T.: 0.007 min
Response: 404709
Conc: 14.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.983 min
Delta R.T.: -0.010 min
Response: 934473
Conc: 20.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.235 min
Delta R.T.: -0.012 min
Response: 441391
Conc: 15.86 ng/ml



#8 AR-1221-1


R.T.: 5.210 min
Delta R.T.: 0.000 min
Response: 55069
Conc: 103.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
NWB-1630




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.207 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.313 min
Delta R.T.: 0.004 min
Response: 28787
Conc: 71.06 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.305 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 5.397 min
Delta R.T.: 0.000 min
Response: 4892751
Conc: 4015.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
NWB-1630




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 4.393 min
Response: 0
Conc: N.D.




#11 AR-1232-1

R.T.: 5.397 min
Delta R.T.: 0.001 min
Response: 4892751
Conc: 4403.47 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T. : 4.393 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 5.969 min
Delta R.T.: -0.016 min
Response: 40497
Conc: 67.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
NWB-1630




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 5.222 min
Response: 0
Conc: N.D.




#15 AR-1232-5

R.T.: 6.572 min
Delta R.T.: -0.003 min
Response: 13184
Conc: 34.28 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T. : 5.694 min
Response: 0
Conc: N.D.



#22 AR-1248-2


R.T.: 6.572 min
Delta R.T.: -0.002 min
Response: 13184
Conc: 8.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
NWB-1630




#22 AR-1248-2

R.T.: 0.000 min
Exp R.T. : 5.465 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 7.216 min
Delta R.T.: -0.004 min
Response: 15745
Conc: 8.33 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 6.592 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 7.200 min
Delta R.T.: 0.011 min
Response: 24036
Conc: 13.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
NWB-1630




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 6.073 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.818 min
Delta R.T.: 0.018 min
Response: 75962
Conc: 24.64 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.652 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 8.096 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
NWB-1630




#29 AR-1254-4

R.T.: 6.866 min
Delta R.T.: -0.027 min
Response: 10000
Conc: 7.37 ng/ml




#31 AR-1260-1

R.T.: 7.974 min
Delta R.T.: 0.007 min
Response: 10854
Conc: 4.93 ng/ml



#31 AR-1260-1


R.T.: 6.775 min
Delta R.T.: -0.016 min
Response: 18817
Conc: 13.62 ng/ml



#32 AR-1260-2


R.T.: 8.207 min
Delta R.T.: -0.027 min
Response: 12364
Conc: 4.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
NWB-1630




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 6.989 min
Response: 0
Conc: N.D.




#35 AR-1260-5

R.T.: 9.149 min
Delta R.T.: -0.009 min
Response: 35869
Conc: 7.77 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T. : 7.876 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 9.149 min
Delta R.T.: -0.008 min
Response: 35869
Conc: 6.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
NWB-1630




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.875 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 9.477 min
Delta R.T.: 0.008 min
Response: 228164
Conc: 59.55 ng/ml



#38 AR-1262-3


R.T.: 8.146 min
Delta R.T.: -0.016 min
Response: 11993
Conc: 8.52 ng/ml



#41 AR-1268-1

R.T.: 9.477 min
Delta R.T.: 0.006 min
Response: 228164
Conc: 37.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
NWB-1630



#41 AR-1268-1

R.T.: 8.146 min
Delta R.T.: -0.017 min
Response: 11993
Conc: 3.01 ng/ml