

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072821\
 Data File : PP037775.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jul 2021 19:08
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 03:55:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.954	3.945	2407967	1408710	52.981	51.480
2)	SA Decachlor...	10.983	9.235	2265659	1209528	49.215	43.454
Target Compounds							
26)	L6 AR-1254-1	7.184	6.064	913763	694747	508.636	467.978
27)	L6 AR-1254-2	7.413	6.225	1309309	578879	470.691	445.633
28)	L6 AR-1254-3	7.795	6.644	1390165	936377	450.910	438.784
29)	L6 AR-1254-4	8.091	6.884	1050988	604009	462.697	445.195
30)	L6 AR-1254-5	8.520	7.313	1140418	827359	455.730	441.344



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072821\
Data File : PP037775.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jul 2021 19:08
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 03:55:46 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.954 min
Delta R.T.: -0.006 min
Response: 2407967
Conc: 52.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.945 min
Delta R.T.: -0.004 min
Response: 1408710
Conc: 51.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.983 min
Delta R.T.: -0.009 min
Response: 2265659
Conc: 49.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.235 min
Delta R.T.: -0.013 min
Response: 1209528
Conc: 43.45 ng/ml



#26 AR-1254-1


R.T.: 7.184 min
Delta R.T.: -0.006 min
Response: 913763
Conc: 508.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.064 min
Delta R.T.: -0.009 min
Response: 694747
Conc: 467.98 ng/ml




#27 AR-1254-2

R.T.: 7.413 min
Delta R.T.: -0.006 min
Response: 1309309
Conc: 470.69 ng/ml



#27 AR-1254-2


R.T.: 6.225 min
Delta R.T.: -0.009 min
Response: 578879
Conc: 445.63 ng/ml



#28 AR-1254-3


R.T.: 7.795 min
Delta R.T.: -0.005 min
Response: 1390165
Conc: 450.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.644 min
Delta R.T.: -0.008 min
Response: 936377
Conc: 438.78 ng/ml




#29 AR-1254-4

R.T.: 8.091 min
Delta R.T.: -0.005 min
Response: 1050988
Conc: 462.70 ng/ml



#29 AR-1254-4


R.T.: 6.884 min
Delta R.T.: -0.009 min
Response: 604009
Conc: 445.19 ng/ml



#30 AR-1254-5

R.T.: 8.520 min
Delta R.T.: -0.005 min
Response: 1140418
Conc: 455.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.313 min
Delta R.T.: -0.009 min
Response: 827359
Conc: 441.34 ng/ml