

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072823\
 Data File : PP059045.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2023 10:26
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 28 22:52:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 22 02:26:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.425	3.662	25385772	39901248	18.440	20.748
2)	SA Decachlor...	10.221	8.703	20532556	32622658	23.834	23.056
Target Compounds							
3)	L1 AR-1016-1	5.596	4.761	47309	36171	1.083	0.629 #
4)	L1 AR-1016-2	5.616	4.774	64667	26979	1.039	0.328 #
5)	L1 AR-1016-3	5.688	4.977f	42427	168812	1.060	3.839 #
6)	L1 AR-1016-4	5.792	5.023f	19683	160508	0.610	4.128 #
7)	L1 AR-1016-5	6.087	5.212	12071	726858	0.354	14.694 #
8)	L2 AR-1221-1	4.637	3.894	16641	178695	0.936	7.571 #
9)	L2 AR-1221-2	4.735	3.964	542510	555165	41.386	31.128
10)	L2 AR-1221-3	4.797	4.041	11246	38186	0.290	0.692 #
11)	L3 AR-1232-1	4.797	4.041	11246	38186	0.350	0.841 #
12)	L3 AR-1232-2	5.330	4.774	12557	26979	0.693	0.699
13)	L3 AR-1232-3	5.616	4.977f	64667	168812	2.183	8.125 #
14)	L3 AR-1232-4	5.792	5.044	19683	131714	1.311	6.508 #
15)	L3 AR-1232-5	5.875	5.212	11256	726858	0.869	32.511 #
16)	L4 AR-1242-1	5.596	4.761	47309	36171	1.251	0.728 #
17)	L4 AR-1242-2	5.616	4.774	64667	26979	1.201	0.378 #
18)	L4 AR-1242-3	5.688	4.977f	42427	168812	1.225	4.452 #
19)	L4 AR-1242-4	5.792	5.044	19683	131714	0.699	3.272 #
20)	L4 AR-1242-5	6.529	5.551f	65007	547634	2.291	11.376 #
21)	L5 AR-1248-1	5.596	4.761	47309	36171	1.710	1.004 #
22)	L5 AR-1248-2	5.875	5.023f	11256	160508	0.262	2.832 #
23)	L5 AR-1248-3	6.087	5.044	12071	131714	0.258	2.300 #
24)	L5 AR-1248-4	6.487	5.212	130539	726858	2.760	10.694 #
25)	L5 AR-1248-5	6.529	5.589f	65007	1366849	1.418	22.430 #
26)	L6 AR-1254-1	6.472	5.589f	96357	1366849	1.817	13.516 #
27)	L6 AR-1254-2	6.689	5.719	55196	749891	0.705	8.674 #
28)	L6 AR-1254-3	7.047	6.124	967568	285401	12.346	2.077 #
29)	L6 AR-1254-4	7.331	6.346	27990	518966	0.536	6.495 #
30)	L6 AR-1254-5	7.761	6.771	106324	43857	1.977	0.364 #
31)	L7 AR-1260-1	7.218	6.266	628927	320858	10.844	3.242 #
32)	L7 AR-1260-2	7.484	6.416f	39780	1292794	0.657	11.371 #
33)	L7 AR-1260-3	7.830	0.000	47962	0	1.083	N.D. #
34)	L7 AR-1260-4	8.063	7.057	64172	64275	1.301	0.756 #
35)	L7 AR-1260-5	8.380	7.320	54426	9360	0.610	0.055 #
36)	L8 AR-1262-1	7.830	6.875	47962	28346	0.703	0.519 #
37)	L8 AR-1262-2	8.380	7.057	54426	64275	0.497	0.557
38)	L8 AR-1262-3	8.686	7.609	31977	36558	0.408	0.415
39)	L8 AR-1262-4	8.793	7.665	37445	33420	0.619	0.212 #
40)	L8 AR-1262-5	9.453	8.181f	12188	93427	0.314	1.371 #
41)	L9 AR-1268-1	8.686	7.609	31977	36558	0.239	0.151 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072823\
Data File : PP059045.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2023 10:26
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 22:52:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 02:26:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.793	7.665	37445	33420	0.315	0.151 #
43)	L9 AR-1268-3	9.028	7.867	39213	125317	0.366	0.615 #
44)	L9 AR-1268-4	9.453	8.181f	12188	93427	0.293	1.264 #
45)	L9 AR-1268-5	9.872	8.445	131747	233340	0.411	0.424



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072823\
Data File : PP059045.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2023 10:26
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 22:52:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 02:26:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.425 min
Delta R.T.: 0.000 min
Response: 25385772
Conc: 18.44 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.662 min
Delta R.T.: -0.002 min
Response: 39901248
Conc: 20.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.221 min
Delta R.T.: -0.005 min
Response: 20532556
Conc: 23.83 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.703 min
Delta R.T.: -0.012 min
Response: 32622658
Conc: 23.06 ng/ml



#3 AR-1016-1


R.T.: 5.596 min
Delta R.T.: -0.006 min
Response: 47309
Conc: 1.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.761 min
Delta R.T.: 0.000 min
Response: 36171
Conc: 0.63 ng/ml




#4 AR-1016-2

R.T.: 5.616 min
Delta R.T.: -0.008 min
Response: 64667
Conc: 1.04 ng/ml



#4 AR-1016-2


R.T.: 4.774 min
Delta R.T.: -0.007 min
Response: 26979
Conc: 0.33 ng/ml



#5 AR-1016-3


R.T.: 5.688 min
Delta R.T.: 0.001 min
Response: 42427
Conc: 1.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.977 min
Delta R.T.: 0.019 min
Response: 168812
Conc: 3.84 ng/ml




#6 AR-1016-4

R.T.: 5.792 min
Delta R.T.: 0.006 min
Response: 19683
Conc: 0.61 ng/ml



#6 AR-1016-4


R.T.: 5.023 min
Delta R.T.: 0.022 min
Response: 160508
Conc: 4.13 ng/ml



#7 AR-1016-5


R.T.: 6.087 min
Delta R.T.: 0.004 min
Response: 12071
Conc: 0.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.212 min
Delta R.T.: -0.002 min
Response: 726858
Conc: 14.69 ng/ml




#8 AR-1221-1

R.T.: 4.637 min
Delta R.T.: 0.003 min
Response: 16641
Conc: 0.94 ng/ml



#8 AR-1221-1


R.T.: 3.894 min
Delta R.T.: 0.012 min
Response: 178695
Conc: 7.57 ng/ml



#9 AR-1221-2


R.T.: 4.735 min
Delta R.T.: 0.014 min
Response: 542510
Conc: 41.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.964 min
Delta R.T.: -0.004 min
Response: 555165
Conc: 31.13 ng/ml




#10 AR-1221-3

R.T.: 4.797 min
Delta R.T.: 0.000 min
Response: 11246
Conc: 0.29 ng/ml



#10 AR-1221-3


R.T.: 4.041 min
Delta R.T.: -0.004 min
Response: 38186
Conc: 0.69 ng/ml



#11 AR-1232-1


R.T.: 4.797 min
Delta R.T.: -0.001 min
Response: 11246
Conc: 0.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.041 min
Delta R.T.: -0.004 min
Response: 38186
Conc: 0.84 ng/ml




#12 AR-1232-2

R.T.: 5.330 min
Delta R.T.: -0.002 min
Response: 12557
Conc: 0.69 ng/ml



#12 AR-1232-2


R.T.: 4.774 min
Delta R.T.: -0.007 min
Response: 26979
Conc: 0.70 ng/ml



#13 AR-1232-3


R.T.: 5.616 min
Delta R.T.: -0.008 min
Response: 64667
Conc: 2.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.977 min
Delta R.T.: 0.019 min
Response: 168812
Conc: 8.13 ng/ml




#14 AR-1232-4

R.T.: 5.792 min
Delta R.T.: 0.006 min
Response: 19683
Conc: 1.31 ng/ml



#14 AR-1232-4


R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 131714
Conc: 6.51 ng/ml



#15 AR-1232-5


R.T.: 5.875 min
Delta R.T.: -0.003 min
Response: 11256
Conc: 0.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.212 min
Delta R.T.: -0.003 min
Response: 726858
Conc: 32.51 ng/ml




#16 AR-1242-1

R.T.: 5.596 min
Delta R.T.: -0.006 min
Response: 47309
Conc: 1.25 ng/ml



#16 AR-1242-1


R.T.: 4.761 min
Delta R.T.: -0.001 min
Response: 36171
Conc: 0.73 ng/ml



#17 AR-1242-2


R.T.: 5.616 min
Delta R.T.: -0.009 min
Response: 64667
Conc: 1.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.774 min
Delta R.T.: -0.007 min
Response: 26979
Conc: 0.38 ng/ml




#18 AR-1242-3

R.T.: 5.688 min
Delta R.T.: 0.000 min
Response: 42427
Conc: 1.23 ng/ml



#18 AR-1242-3


R.T.: 4.977 min
Delta R.T.: 0.019 min
Response: 168812
Conc: 4.45 ng/ml



#19 AR-1242-4


R.T.: 5.792 min
Delta R.T.: 0.006 min
Response: 19683
Conc: 0.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 131714
Conc: 3.27 ng/ml




#20 AR-1242-5

R.T.: 6.529 min
Delta R.T.: 0.002 min
Response: 65007
Conc: 2.29 ng/ml



#20 AR-1242-5


R.T.: 5.551 min
Delta R.T.: -0.019 min
Response: 547634
Conc: 11.38 ng/ml



#21 AR-1248-1


R.T.: 5.596 min
Delta R.T.: -0.007 min
Response: 47309
Conc: 1.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.761 min
Delta R.T.: -0.002 min
Response: 36171
Conc: 1.00 ng/ml




#22 AR-1248-2

R.T.: 5.875 min
Delta R.T.: -0.004 min
Response: 11256
Conc: 0.26 ng/ml



#22 AR-1248-2


R.T.: 5.023 min
Delta R.T.: 0.022 min
Response: 160508
Conc: 2.83 ng/ml



#23 AR-1248-3


R.T.: 6.087 min
Delta R.T.: 0.003 min
Response: 12071
Conc: 0.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 131714
Conc: 2.30 ng/ml




#24 AR-1248-4

R.T.: 6.487 min
Delta R.T.: -0.002 min
Response: 130539
Conc: 2.76 ng/ml



#24 AR-1248-4


R.T.: 5.212 min
Delta R.T.: -0.004 min
Response: 726858
Conc: 10.69 ng/ml



#25 AR-1248-5


R.T.: 6.529 min
Delta R.T.: 0.000 min
Response: 65007
Conc: 1.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.589 min
Delta R.T.: -0.023 min
Response: 1366849
Conc: 22.43 ng/ml




#26 AR-1254-1

R.T.: 6.472 min
Delta R.T.: 0.009 min
Response: 96357
Conc: 1.82 ng/ml



#26 AR-1254-1


R.T.: 5.589 min
Delta R.T.: 0.019 min
Response: 1366849
Conc: 13.52 ng/ml



#27 AR-1254-2


R.T.: 6.689 min
Delta R.T.: 0.006 min
Response: 55196
Conc: 0.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.719 min
Delta R.T.: 0.000 min
Response: 749891
Conc: 8.67 ng/ml




#28 AR-1254-3

R.T.: 7.047 min
Delta R.T.: -0.003 min
Response: 967568
Conc: 12.35 ng/ml



#28 AR-1254-3


R.T.: 6.124 min
Delta R.T.: 0.000 min
Response: 285401
Conc: 2.08 ng/ml



#29 AR-1254-4


R.T.: 7.331 min
Delta R.T.: -0.005 min
Response: 27990
Conc: 0.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.346 min
Delta R.T.: -0.008 min
Response: 518966
Conc: 6.50 ng/ml




#30 AR-1254-5

R.T.: 7.761 min
Delta R.T.: 0.004 min
Response: 106324
Conc: 1.98 ng/ml



#30 AR-1254-5


R.T.: 6.771 min
Delta R.T.: -0.003 min
Response: 43857
Conc: 0.36 ng/ml



#31 AR-1260-1


R.T.: 7.218 min
Delta R.T.: 0.004 min
Response: 628927
Conc: 10.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.266 min
Delta R.T.: 0.011 min
Response: 320858
Conc: 3.24 ng/ml




#32 AR-1260-2

R.T.: 7.484 min
Delta R.T.: 0.012 min
Response: 39780
Conc: 0.66 ng/ml



#32 AR-1260-2


R.T.: 6.416 min
Delta R.T.: -0.028 min
Response: 1292794
Conc: 11.37 ng/ml



#33 AR-1260-3


R.T.: 7.830 min
Delta R.T.: -0.003 min
Response: 47962
Conc: 1.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.599 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 8.063 min
Delta R.T.: 0.003 min
Response: 64172
Conc: 1.30 ng/ml



#34 AR-1260-4


R.T.: 7.057 min
Delta R.T.: -0.016 min
Response: 64275
Conc: 0.76 ng/ml



#35 AR-1260-5


R.T.: 8.380 min
Delta R.T.: 0.000 min
Response: 54426
Conc: 0.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.320 min
Delta R.T.: 0.005 min
Response: 9360
Conc: 0.06 ng/ml




#36 AR-1262-1

R.T.: 7.830 min
Delta R.T.: -0.003 min
Response: 47962
Conc: 0.70 ng/ml



#36 AR-1262-1


R.T.: 6.875 min
Delta R.T.: 0.010 min
Response: 28346
Conc: 0.52 ng/ml



#37 AR-1262-2


R.T.: 8.380 min
Delta R.T.: 0.000 min
Response: 54426
Conc: 0.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.057 min
Delta R.T.: -0.015 min
Response: 64275
Conc: 0.56 ng/ml




#38 AR-1262-3

R.T.: 8.686 min
Delta R.T.: -0.011 min
Response: 31977
Conc: 0.41 ng/ml



#38 AR-1262-3


R.T.: 7.609 min
Delta R.T.: 0.008 min
Response: 36558
Conc: 0.42 ng/ml



#39 AR-1262-4


R.T.: 8.793 min
Delta R.T.: 0.005 min
Response: 37445
Conc: 0.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.665 min
Delta R.T.: 0.002 min
Response: 33420
Conc: 0.21 ng/ml




#40 AR-1262-5

R.T.: 9.453 min
Delta R.T.: 0.000 min
Response: 12188
Conc: 0.31 ng/ml



#40 AR-1262-5


R.T.: 8.181 min
Delta R.T.: 0.020 min
Response: 93427
Conc: 1.37 ng/ml



#41 AR-1268-1


R.T.: 8.686 min
Delta R.T.: -0.009 min
Response: 31977
Conc: 0.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.609 min
Delta R.T.: 0.007 min
Response: 36558
Conc: 0.15 ng/ml




#42 AR-1268-2

R.T.: 8.793 min
Delta R.T.: 0.003 min
Response: 37445
Conc: 0.31 ng/ml



#42 AR-1268-2


R.T.: 7.665 min
Delta R.T.: 0.000 min
Response: 33420
Conc: 0.15 ng/ml



#43 AR-1268-3


R.T.: 9.028 min
Delta R.T.: 0.004 min
Response: 39213
Conc: 0.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.867 min
Delta R.T.: -0.007 min
Response: 125317
Conc: 0.61 ng/ml




#44 AR-1268-4

R.T.: 9.453 min
Delta R.T.: 0.000 min
Response: 12188
Conc: 0.29 ng/ml



#44 AR-1268-4


R.T.: 8.181 min
Delta R.T.: 0.019 min
Response: 93427
Conc: 1.26 ng/ml



#45 AR-1268-5

R.T.: 9.872 min
Delta R.T.: -0.006 min
Response: 131747
Conc: 0.41 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.445 min
Delta R.T.: -0.012 min
Response: 233340
Conc: 0.42 ng/ml