

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072823\
 Data File : PP059057.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2023 14:26
 Operator : YP\AJ
 Sample : 03819-02 10X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 370

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/31/2023
 Supervised By :Ankita Jodhani 07/31/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 28 23:02:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 22 02:26:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.423	3.661	2233425	5040202	1.622m	2.621 #
2) SA Decachlor...	10.220	8.703	1596642	1923454	1.853	1.359m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072823\
Data File : PP059057.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2023 14:26
Operator : YP\AJ
Sample : 03819-02 10X
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
370


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/31/2023
Supervised By :Ankita Jodhani 07/31/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 23:02:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 02:26:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.423 min
Delta R.T.: -0.002 min
Response: 2233425
Conc: 1.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
370


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/31/2023
Supervised By :Ankita Jodhani 07/31/2023




#1 Tetrachloro-m-xylene

R.T.: 3.661 min
Delta R.T.: -0.003 min
Response: 5040202
Conc: 2.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: -0.006 min
Response: 1596642
Conc: 1.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.703 min
Delta R.T.: -0.011 min
Response: 1923454
Conc: 1.36 ng/ml m