Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072921\
Data File : PP037818.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Jul 2021 20:52
Operator : DD\AJ
Sample : M2940-08
Misc : AR1242 LOQ 50 PPB
ALS Vial : 39 Sample Multiplier: 1

Instrument : ECD_P
ClientSampleId : LOQ-WATER-02-QT3-2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 02:16:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.953	3.944	977113	550255	21.499	20.109
2)	SA Decachlor...	10.982	9.233	970186	537021	21.075	19.293
	Target Compounds						
16)	L4 AR-1242-1	6.273	5.193	117214	56287	91.955	74.681
17)	L4 AR-1242-2	6.297	5.213	159219	81139	87.815	78.855
18)	L4 AR-1242-3	6.361	5.403	108476	41320	94.313	72.564
19)	L4 AR-1242-4	6.471	5.500	79903	40850	83.467	79.262
20)	L4 AR-1242-5	7.253	6.064	97570	54433	93.821	79.969

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072921\
Data File : PP037818.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Jul 2021 20:52
Operator : DD\AJ
Sample : M2940-08
Misc : AR1242 LOQ 50 PPB
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOQ-WATER-02-QT3-2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 02:16:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm