

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072924\
 Data File : PP066512.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jul 2024 00:38
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/31/2024
 Supervised By :Ankita Jodhani 07/31/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 31 08:57:35 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP073024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 31 08:55:38 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.778	4.017	196.7E6	199.9E6	97.007	95.253
2) SA Decachlor...	10.772	9.192	396.1E6	249.6E6	91.573	94.186
Target Compounds						
41) L9 AR-1268-1	9.116	7.985	327.6E6	236.9E6	937.362m	962.226m
42) L9 AR-1268-2	9.221	8.050	305.0E6	217.2E6	924.388	955.023
43) L9 AR-1268-3	9.472	8.268	275.2E6	187.3E6	921.334	951.789
44) L9 AR-1268-4	9.935	8.577	104.9E6	68300864	942.218	941.543
45) L9 AR-1268-5	10.395	8.903	846.7E6	539.0E6	937.153	964.730

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072924\
Data File : PP066512.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2024 00:38
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/31/2024
Supervised By :Ankita Jodhani 07/31/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 31 08:57:35 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP073024.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 31 08:55:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.778 min
Delta R.T.: 0.000 min
Response: 196695027
Conc: 97.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/31/2024
Supervised By :Ankita Jodhani 07/31/2024




#1 Tetrachloro-m-xylene

R.T.: 4.017 min
Delta R.T.: 0.000 min
Response: 199946251
Conc: 95.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.772 min
Delta R.T.: 0.001 min
Response: 396104557
Conc: 91.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.192 min
Delta R.T.: 0.000 min
Response: 249594807
Conc: 94.19 ng/ml




#41 AR-1268-1

R.T.: 9.116 min
Delta R.T.: 0.000 min
Response: 327550534
Conc: 937.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/31/2024
Supervised By :Ankita Jodhani 07/31/2024




#41 AR-1268-1

R.T.: 7.985 min
Delta R.T.: 0.000 min
Response: 236944018
Conc: 962.23 ng/ml




#42 AR-1268-2

R.T.: 9.221 min
Delta R.T.: 0.001 min
Response: 305034453
Conc: 924.39 ng/ml



#42 AR-1268-2

R.T.: 8.050 min
Delta R.T.: 0.000 min
Response: 217238220
Conc: 955.02 ng/ml




#43 AR-1268-3

R.T.: 9.472 min
Delta R.T.: 0.000 min
Response: 275195668
Conc: 921.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/31/2024
Supervised By :Ankita Jodhani 07/31/2024




#43 AR-1268-3

R.T.: 8.268 min
Delta R.T.: 0.000 min
Response: 187264219
Conc: 951.79 ng/ml




#44 AR-1268-4

R.T.: 9.935 min
Delta R.T.: 0.000 min
Response: 104886931
Conc: 942.22 ng/ml



#44 AR-1268-4

R.T.: 8.577 min
Delta R.T.: 0.000 min
Response: 68300864
Conc: 941.54 ng/ml




#45 AR-1268-5

R.T.: 10.395 min
Delta R.T.: 0.000 min
Response: 846684178
Conc: 937.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/31/2024
Supervised By :Ankita Jodhani 07/31/2024



#45 AR-1268-5

R.T.: 8.903 min
Delta R.T.: 0.000 min
Response: 539009882
Conc: 964.73 ng/ml