

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP073018.M
 Title : 8080.M
 Last Update : Tue Jul 31 01:31:14 2018
 Response Via : Initial Calibration

Calibration Files

750 =PP026907.D 1000 =PP026908.D 1500 =PP026909.D
 500 =PP026906.D 200 =PP026905.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.900	5.704	5.518	5.945	6.309	5.875	E5	5.04
2) T	3,5-DICHLOROBENZOI...	4.682	4.505	4.336	4.645	4.847	4.603	E5	4.19
3) T	4-Nitrophenol	4.696	4.566	4.463	4.577	4.632	4.587	E5	1.89
4) S	2,4-DCAA	2.434	2.358	2.300	2.386	2.463	2.388	E5	2.67
5) T	DICAMBA	1.433	1.394	1.345	1.413	1.477	1.413	E6	3.47
6) T	MCPP	4.338	4.305	4.175	3.906	2.673	3.880	E5	17.93
7) T	MCPA	7.417	7.550	7.735	6.231	4.620	6.710	E5	19.50
8) T	DICHLORPROP	3.227	3.125	2.997	3.232	3.221	3.161	E5	3.21
9) T	2,4-D	3.976	3.865	3.720	3.903	4.072	3.907	E5	3.35
10) T	Pentachlorophenol	7.771	7.328	6.779	7.896	8.281	7.611	E6	7.57
11) T	2,4,5-TP (SILVEX)	2.914	2.792	2.655	2.905	3.079	2.869	E6	5.48
12) T	2,4,5-T	2.593	2.518	2.386	2.566	2.695	2.552	E6	4.43
13) T	2,4-DB	2.304	2.174	2.100	2.164	2.192	2.187	E5	3.39
14) T	DINOSEB	2.072	1.991	1.894	2.031	2.125	2.022	E6	4.32
15) T	Picloram	4.262	4.297	3.991	4.100	4.000	4.130	E6	3.48
16) T	DCPA	4.068	3.921	3.701	4.111	4.256	4.011	E6	5.25

Signal #2 Calibration Files

750 =PP026907.D 1000 =PP026908.D 1500 =PP026909.D
 500 =PP026906.D 200 =PP026905.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.653	5.454	5.261	5.649	6.075	5.618	E5	5.38
2) T	3,5-DICHLOROBENZOI...	5.439	5.241	5.076	5.398	5.597	5.350	E5	3.71
3) T	4-Nitrophenol	2.494	2.420	2.383	2.414	2.401	2.422	E5	1.75
4) S	2,4-DCAA	4.095	3.952	3.830	4.052	4.318	4.049	E5	4.48
5) T	DICAMBA	1.653	1.586	1.510	1.647	1.741	1.627	E6	5.29
6) T	MCPP	9.390	9.354	9.381	8.541	6.423	8.618	E5	14.84
7) T	MCPA	1.384	1.386	1.400	1.235	0.884	1.258	E6	17.45
8) T	DICHLORPROP	3.912	3.792	3.632	3.912	4.159	3.881	E5	4.98
9) T	2,4-D	5.364	5.171	4.944	5.282	5.610	5.274	E5	4.65
10) T	Pentachlorophenol	6.742	6.266	5.662	6.992	7.690	6.671	E6	11.44
11) T	2,4,5-TP (SILVEX)	2.633	2.497	2.347	2.663	2.836	2.595	E6	7.08
12) T	2,4,5-T	2.214	2.144	2.001	2.221	2.312	2.179	E6	5.31
13) T	2,4-DB	2.926	2.837	2.745	2.835	2.905	2.850	E5	2.50
14) T	DINOSEB	1.845	1.759	1.656	1.839	1.925	1.805	E6	5.65
15) T	Picloram	3.545	3.513	3.223	3.482	3.482	3.449	E6	3.74
16) T	DCPA	3.257	3.060	2.837	3.332	3.607	3.218	E6	9.00

(#) = Out of Range