

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080122\
 Data File : PP050038.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2022 22:03
 Operator : YP\AJ
 Sample : N3983-01 10X
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 03:00:53 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 06:08:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.350	3.598	2738281	4810902	1.887	1.968
2)	SA Decachlor...	10.104	8.565	2382226	1814095	1.684	1.558
Target Compounds							
6)	L1 AR-1016-4	0.000	4.909	0	48969	N.D.	0.921 #
20)	L4 AR-1242-5	0.000	5.489f	0	926487	N.D.	14.575 #
22)	L5 AR-1248-2	0.000	4.909	0	48969	N.D.	0.658 #
25)	L5 AR-1248-5	0.000	5.489f	0	926487	N.D.	11.116 #
26)	L6 AR-1254-1	0.000	5.489f	0	926487	N.D.	7.305 #
32)	L7 AR-1260-2	0.000	6.318f	0	200021	N.D.	1.662 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080122\
Data File : PP050038.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2022 22:03
Operator : YP\AJ
Sample : N3983-01 10X
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 03:00:53 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 06:08:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.350 min
Delta R.T.: -0.003 min
Response: 2738281
Conc: 1.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.598 min
Delta R.T.: -0.004 min
Response: 4810902
Conc: 1.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.104 min
Delta R.T.: -0.004 min
Response: 2382226
Conc: 1.68 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.565 min
Delta R.T.: -0.008 min
Response: 1814095
Conc: 1.56 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T.: 5.715 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#6 AR-1016-4

R.T.: 4.909 min
Delta R.T.: -0.002 min
Response: 48969
Conc: 0.92 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T.: 6.453 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.489 min
Delta R.T.: 0.018 min
Response: 926487
Conc: 14.58 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 5.801 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#22 AR-1248-2

R.T.: 4.909 min
Delta R.T.: 0.000 min
Response: 48969
Conc: 0.66 ng/ml




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T.: 6.453 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 5.489 min
Delta R.T.: -0.023 min
Response: 926487
Conc: 11.12 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T.: 6.385 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 5.489 min
Delta R.T.: 0.018 min
Response: 926487
Conc: 7.30 ng/ml



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T.: 7.395 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 6.318 min
Delta R.T.: -0.019 min
Response: 200021
Conc: 1.66 ng/ml